



**VIEW LAB MONTHLY SUMMARY**

<b>Lab Name</b>	Govt. Block PHC, Sirukambur	<a href="#">Details About Robust Analysis</a>
<b>Lab No</b>	7238	<a href="#">Detail About Monthly Summary</a>
<b>Month</b>	August	<a href="#">Detail about SDI</a>
<b>Year</b>	2023	
<b>Constituent Group</b>	Chemistry III	

[Click on the analyte to view Graphical Data](#)

Date of Result Entered : 16/08/2023

Date of Report Published : 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1841	175.58	5.13	9.00	182 mg/dL	0.71	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.76	6.77	0.12	2.1 mg/dL	2.86	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1764	113.80	4.07	4.64	113 mg/dL	-0.17	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>