



CALIBRATION CERTIFICATE

CERTIFICATE NO: SBS/CL/23/15414		MEDICAL DEVICES		Page No:1 of 1
Issue Date	18-10-2023			
SRF No & Date	SRF/23/00914-0003 & 17-10-2023			
Receipt Date	17-10-2023			
Calibration Date	17-10-2023			
Calibration Due	16-10-2024			
Customer Name & Address				
GOVERNMENT URBAN PRIMARY HEALTH CENTRE, ANUMANTHAPUTHERI-603002,CHENGALPATTU DISTRICT.				
Details of Device Under Calibration (DUC)				
Description	: SEMI AUTO ANALYZER	Make & Model	: ROBONIK & PRIETEST TOUCH	
Range	: MULTI	Sr. No	: ATCD0271220RBK	
Resolution	: MULTI	Identification No	: NA	
DUC Condition	: SATISFACTORY	Location	: LABORATORY	
Environmental Conditions & Calibration Procedure Details				
Environmental Details	Temperature:25.6°C	Relative Humidity	52% RH	
Calibration Procedure No	SBS/CP/MD/20	Calibration done at	ONSITE	
Reference Standards Details				
S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	M-230809-16-4	10-08-2024

ELECTRICAL SAFETY


RESULTS

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resistance >20MΩ	Measured values in MΩ 93	Uncertainty in % (±) 13.92
2	Earth Leakage <5000μAfor B,BF,CF	Measured values in μA 177	Uncertainty in % (±) 9.8
3	Enclosure Leakage <500μAfor B,BF,CF	Measured values in μA 216	Uncertainty in % (±) 8.8

REMARKS

- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2.The user should determine the suitability of the instrument for its intended use.
- 3.The recalibration interval should be determined on the user requirement.
- 4.The results stated in this certificate relate only to the item calibrated.
- 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By


(Calibration Engineer)
P.PRASANNA



Authorised Signatory


Technical Manager
(C.SHANMUGARAJ)

Chief Executive