




CALIBRATION / TESTING REPORT							
CUSTOMER DETAILS :			CERTIFICATE NO	TICS/2023/VP/SG/04			
			GOVERNMENT PRIMARY HEALTH CENTRE V.PUDUR, SIVAGANGAI.			EQUIPMENT NAME	MICROPIPETTE
CALIBRATION DATE			20/10/2023	MANUFACTURE	THERMO SCIENTIFIC		
			CALIBRATION DUE DATE			19/10/2024	MODEL
			ASSET NO	VP PH 04			
			INSTALLED DEPARTMENT	LAB			
1. Have more dust in machine outside <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No 2. Any Physical damaged in machine <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No 3. Machine inside Cleaning (if required) <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No							
TEST EQUIPMENT	ID. NO	CERTIFICATE NUMBER	VALIDITY				
ELECTRONIC SEMI MICRO BALANCE	NCS/WB01	TSC/23-24/5424-1	27-JUNE-2024				
TEST REPORT					Temperature 25°C Humidity RH 55%		
PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS		
VOLUME	µL	10	10.02	± 1	PASS		
		100	99.99	± 1	PASS		
*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipments having traceability to National/International standards.							
<input checked="" type="checkbox"/> ACCEPTABLE FOR USE <input type="checkbox"/> CORRECTIVE MAINTENANCE REQUIRED <input type="checkbox"/> REMOVED FROM USE							
Calibrated By :				Authorized Sign With Seal			
							
							

CALIBRATION / TESTING REPORT					
CUSTOMER DETAILS :			CERTIFICATE NO	TICS/2023/VP/SG/05	
			EQUIPMENT NAME	MICROPIPETTE	
GOVERNMENT PRIMARY HEALTH CENTRE V.PUDUR, SIVAGANGAI.			MANUFACTURE	THERMO SCIENTIFIC	
			MODEL	FINE PIPETTE 100-1000 µL	
CALIBRATION DATE	20/10/2023	ASSET NO	VP PH 05		
CALIBRATION DUE DATE	19/10/2024	INSTALLED DEPARTMENT	LAB		
<p>1. Have more dust in machine outside <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No</p> <p>2. Any Physical damaged in machine <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No</p> <p>3. Machine inside Cleaning (if required) <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No</p>					
TEST EQUIPMENT	ID. NO	CERTIFICATE NUMBER	VALIDITY		
ELECTRONIC SEMI MICRO BALANCE	NCS/WB01	TSC/23-24/5424-1	27-JUNE-2024		
TEST REPORT					Temperature 25°C Humidity RH 55%
PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
VOLUME	µL	100	100.02	± 1	PASS
		1000	1000.08	± 1	PASS
*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipments having traceability to National/International standards.					
<input checked="" type="checkbox"/> ACCEPTABLE FOR USE <input type="checkbox"/> CORRECTIVE MAINTENANCE REQUIRED <input type="checkbox"/> REMOVED FROM USE					
Calibrated By :			Authorized Sign With Seal		
