

G MULTI SERVICE

78, Yahoob Street, Gobichettipalayam, Erode Dist, Tamil Nadu. Pin:638 452. gmultiservicesgobi@gmail.com. Cell: 96597 20654. **Bio Medical Instruments Services**

Er. A. Ibathul Raziq DECE., BE (ICE) Bio Medical Engineer

Calibration Protocol & Report for Medical Equipment

Equipment: PIPETTE

MAKE : FINNEPIPETTE

MODEL :100-1000µL PROTECT CLASS / TYPE : Risk Class II

S. No.

: KA/PI/001

Location: GPHC, KAVERIPATTINAM

S.No	<i>N</i>	Measurement Parameters	Status	Remarks
1	Electrical Test	(a)Safety Standards (range: 210-240 V) (b)Equipotential ground & Earthing (c)Risk of shock/fire Hazards,if any (d)Chassis & Insulation Resistance (e) Earth & Chassis Leakage Current (f) Battery conditions & charging status		-
2	Physical Condition Test	(a)Cabinet & Visible damages, if any (b) Risk of mechanical damages, if any	No No	Checked OK
3	Performance Test	(e) Mean 99.2μl (f) A -0.5 μl (g) Acc % -0.71 (h) CV% -0.36	OK OK OK	
4	Calibration Test	min volume 100 μl max Volume 1000 μl (ELECTRONIC BALANCEING TEST) (e) A+or - 1.5 , 5 μl (f) S 1.5 μl max Vol. 1000 μl (g) Acc+ or - 1.50, 0.50 (h) CV% 0.50,0.10	Ok Ok Ok Ok	Tested
Formulated By: Edition Research Laboratory Endorsed By: Research team head Reviewed On: 20.10.23 Next Review: 19.10.24		Issued By:	User Department:	