



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - APRIL 2023

13117
 19/04/2023
 06/05/2023

Lab Name: Govt PHC, Seerpanandal
 Constituent Group: Chemistry III
 PT item: Lyophilized human serum based

Lab No:
 Date of Result Entered:
 Date of Report Published:

Sl. No.	Parameter	Method	Instrument	Result	Target	Delta	Delta 2	Delta 3	Delta 4	Delta 5	Delta 6	Delta 7	Delta 8	Delta 9	Delta 10
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	119 mg/dL	-0.06	0.34					
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	1.8 mg/dL	-2.06	0.01					
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	110 mg/dL	-0.28	0.28					

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS .
 Govt. PHC, Seerpanandal
 Seerpanandal
 TAMILNADU

Coordinator Contact Details.
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pameia Christudoss
 Dr. Pameia Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential!
 CMC EQAS does not sub contract any components
 ***** End of Report *****

G. Ravi