



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME



PC-1024

MONTHLY SUMMARY REPORT - MAY 2023

Lab Name Govt. PHC, Seerpanandal
Constituent Group Chemistry III
PT item Lyophilized human serum based

Lab No 13117
Date of Result Entered : 19/05/2023
Date of Report Published : 05/06/2023

Sl. No.	Name of Analyte	Method	Instrument	Sample No.	Result	Target	Delta	Delta Z	Delta X	Delta Y
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1834	180.88	5.19	9.39	169 mg/dL	-1.26	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1854	1.76	7.48	0.13	1.7 mg/dL	-0.45	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.83	112 mg/dL	-0.45	0.27

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
Govt. PHC, Seerpanandal
Seerpanandal
TAMILNADU

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

G. Devi
MEDICAL OFFICER,
Govt. Primary Health Centre,
Seerpanandal 605802
Kallakurichi Dist