



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JUNE 2023



Lab Name: Govt. PHC, Seerpanandal
 Constituent Group: Chemistry III
 PT item: Lyophilized human serum based

Lab No: 13117
 Date of Result Entered: 19/06/2023
 Date of Report Published: 04/07/2023

Sl. No.	Test Name	Method	Instrument	No. of Patients	Mean	SD	CV (%)	Target	SDI	SI
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	272 mg/dL	-0.99	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.3 mg/dL	-1.58	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	104 mg/dL	-1.92	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
 Govt. PHC, Seerpanandal
 Seerpanandal
 TAMILNADU

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****

G. Ravi Var
 MEDICAL OFFICER,
 Govt. Primary Health Centre,
 Seerpanandal 605802
 Kallakurichi Dist