

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Seerpanandal  
 Lab No 13117  
 Month July  
 Year 2023  
 Constituent Group Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 20/07/2023

Date of Report Published : 07/08/2023

Sl. No	Analyte	Method	Instrument	Count	Mean	SD	CV	Target	SDI	Delta
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1876	109.35	5.55	6.07	109 mg/dL	-0.06	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1850	5.83	6.00	0.35	5.2 mg/dL	-1.80	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1776	113.00	4.12	4.66	112 mg/dL	-0.21	0.22

SDI Range	Interpretation
Within -1.00 to -1.00	Excellent.
Within +1.01 to -2.00	Good.
Within +2.01 to +2.99	Accept with caution. Warning Signal.
Beyond +3.0	Unacceptable performance. Action Signal.

Links

MEDICAL OFFICER,  
 Govt Primary Health Centre,  
 Seerpanandal, Taluk,  
 Kallakurichi Dist.