



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - AUGUST 2023



Lab Name Govt. PHC, Seerpanandal  
Constituent Group Chemistry III  
PT item Lyophilized human serum based

Lab No 13117  
Date of Result Entered : 20/08/2023  
Date of Report Published : 01/09/2023

Sl. No.	Parameter	Method	Instrument	Target	Result	Delta	Delta 2	Delta 3	Delta 4	Delta 5
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1841	175.58	5.13	9.00	173 mg/dL	-0.29	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.76	6.77	0.12	1.9 mg/dL	1.18	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1764	113.80	4.07	4.64	113 mg/dL	-0.17	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

LAB ADDRESS :  
Govt. PHC, Seerpanandal  
Seerpanandal  
TAMILNADU

Coordinator Contact Details:  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact Number: 0416-2283102

**MEDICAL OFFICER,**  
Govt. Primary Health Centre,  
Seerpanandal 605802  
Kallakurichi Dist

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential