



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - SEPTEMBER 2023



PC-1024

Lab Name: Govt. PHC, Seerpanandal Lab No: 13117
 Constituent Group: Chemistry III Date of Result Entered: 20/09/2023
 PT Item: Lyophilized human serum based Date of Report Published: 05/10/2023

Sl No	Parameter	Method / Principle Name	Instrument Name	No. of Participants	CV	CV	CV	CV	CV	CV
1	GLUCOSE III	GOD-POD III	Robcnik (Semi automatic / Fully Automatic)	1870	310.86	4.82	14.98	312 mg/dL	0.08	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robcnik (Semi automatic / Fully Automatic)	1852	1.29	8.14	0.10	1.1 mg/dL	-1.81	0.00
3	CHOLESTROL III	CHOD-PAP III	Robcnik (Semi automatic / Fully Automatic)	1783	113.62	3.80	4.32	112 mg/dL	-0.38	0.20

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. PHC, Seerpanandal

Seerpanandal

TAMILNADU

Coordinator Contact Details
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential