



## G MULTI SERVICES


78, Yashob Street, Gobichettipalayam, Erode Dist, Tamil Nadu. Pin: 638 452.  
gmultiservicesgobin@gmail.com. Cell : 96597 20664.  
Bio Medical Instruments Services

Er. A. Ipathul Razip DECE., BE (ICE)  
Bio Medical Engineer

### Calibration Protocol & Report for Medical Equipment

EQUIPMENT : PIPETTE  
MAKE : FINNPIPEF3  
MODEL : 10-100 µl

PROTECT CLASS/TEST : Risk class II  
S No : 4640040-N  
Location : PHC Susaipuram

S.no	Measurement Parameters	Status	Remarks
1	<b>Electrical Test</b> (a) Safety Standards (range : 210-240v) (b) Equipotential ground & Earthing (c) Risk of shock/fire Hazards, if any (d) Chassis & Insulation Resistance (e) Earth & Chassis Leakage current (f) Battery conditions & charging status	----	-
2	<b>Physical Condition Test</b> (a) Cabinet & Visible damages, if any (b) Risk of mechanical damages, if any	No No	Checked Ok
3	<b>Performance Test</b> (a) Mean 99.2 µl (b) A -0.5 µl (c) Acc % -0.71 (d) CV% -0.36	Ok Ok Ok Ok	
4	<b>Calibration Test</b> (a) Min volume 10 µl (b) Max volume 100 µl  (c) A+or - 1.5, 5 µl (d) S 1.5 ul max vol. 10 µl (e) Acc+or -1.50, 0.50 (f) CV% 0.50, 0.20		Tested  Ok Ok Ok Ok
Formulated By: Edition Research Laboratory		Issued By:	User Department:
Endorsed By: Research Team Head			
Reviewed On: 01.10.23 Next Review :29.09.24			