

G MULTI SERVICES
78, Yahoob Bireet, Gobichettipaleyam, Erede Dist, Tamili Nadu. Pin 638 452.
gmultiservicesgobing mail soom. Cell: 94597 20454.
Bio Medical Instruments Services

Er. A. Ibathul Razip DECE.,BE (ICE) Blo Medical Engineer

Calibration Protocol & Report for Medical Equipment

EQUIPMENT: CENTRIFUGE MAKE: REMI MODEL: R & C with R81

PROTECT CLASS/TEST : Risk class II S.No : -

Location.

: PHC Susaipuram

5.no	1	Measurement Parameters	Status	Remarks
1	Electrical Test	(a) Safety Standards (range :210-240v) (b) Equipotential ground & Earthing (c) Risk of shock/fire Hazards, if any (d) Chassis & Insulation Resistance (e) Earth & Chassis Leakage current (f) Battery conditions & charging status	235 volts 0 to 5 volts No Zero ohms 10MA	Stabilizer/ Ups With in Rangs Well
2	Physical Condition Test	(a) Cabinet & Visible damages, if any (b) Front panet/Top lid, buckets, motor assembly & power cord (c) Risk of mechanical damages, if any	No OK	Checked Ok
3	Performance Test Calibration Test	(a) On/Off switch (b) Speed regulator switch & functional test (c) Running test (d) Abnormalities (e) Display checkup (f) Keypad/indicators & connector (a) Vibration test (b) Speed regulation gradual hike/wobbling test (c) Running test with load (d) Display check up	NO Normal Ok Accurate Ok No Well Well Done Ok	Ok Perfect Perfect
Formulated By: Edition Research Laboratory		(e) Timer test Issued By:	User Department:	
	By: Team On: 01.10.23 ew : 29.09.24	COR.		