



CALIBRATION CERTIFICATE


CERTIFICATE NO:	SBS/CL/23/06573	MECHANICAL	Page No:1 of 1	
Issue Date	02-06-2023			
SRF No & Date	SRF/23/00210-0003 & 01-06-2023			
Receipt Date	01-06-2023			
Calibration Date	01-06-2023			
Calibration Due	31-05-2024			
Customer Name & Address				
GOVERNMENT URBAN PRIMARY HEALTH CENTRE, MARIYANATHAPURAM-624003,DINDIGUL DISTRICT.				
Details of Device Under Calibration (DUC)				
Description	: CENTRIFUGE	Make & Model	: REMI & NA	
Range	: 3500 RPM	Sr. No	: NA	
Least Count	: 10 RPM	Identification No	:	
DUC Condition	: Satisfactory	Location	: LABORATORY	
Environmental Conditions & Standard Operating Procedure Details				
Environmental Details	Temperature: 25.3 °C	Relative Humidity	45 % Rh	
Calibration Procedure No	SBS/CP/ML/04	Calibration done at	ONSITE	
Reference Standards Details				
S.No	Description	Make/ SI No:	Certificate No	Validity
1	Digital Tachometer	LINE SEIKI / 175-0034V	TMS/22/63 01	10-11-2023

CALIBRATION RESULTS

S.No	DEVICE UNDER CALIBRATION	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY (±)
	RPM	RPM	RPM	%
1	1000	999.6	0.4	4.2
2	2000	1999.3	0.7	4.2
3	3500	3499.6	0.4	4.2

REMARKS

1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
2. The user should determine the suitability of the instrument for its intended use.
3. The recalibration interval should be determined on the user requirement.
4. The results stated in this certificate relate only to the item calibrated.
5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

 (Calibration Engineer)
 C.SIVABALAN



Quality Manager
 D.VETRISELVI

Authorised Signatory
 Chief Executive