



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - SEPTEMBER 2023



PC-1024

Lab Name **JEEVAN DIAGNOSTIC CENTRE** Lab No **11850**  
 Constituent Group **Chemistry II** Date of Result Entered : **20/09/2023**  
 PT item **Lyophilized human serum based** Date of Report Published : **05/10/2023**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	1697	333.34	11.48	38.25	335.25 mg/dL	0.05	1.86
2	UREA II	Urease UV / GLDH II	Any Analyser (Automation / Semi Automation)	1143	38.62	12.79	4.94	39.8 mg/dL	0.24	0.29
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Any Analyser (Automation / Semi Automation)	1241	1.16	14.35	0.17	1.15 mg/dL	-0.06	0.01
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Any Analyser (Automation / Semi Automation)	1395	5.18	15.03	0.78	5.31 mg/dL	0.17	0.04
5	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	1299	5.64	10.51	0.59	5.23 g/dL	-0.69	0.03
6	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	1246	3.31	9.55	0.32	3.1 g/dL	-0.66	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	1315	9.82	13.19	1.30	10.65 mg/dL	0.64	0.07
8	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	1501	119.45	10.95	13.08	136.52 mg/dL	1.31	0.67
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	1333	201.48	11.49	23.15	220.65 mg/dL	0.83	1.27

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

**JEEVAN DIAGNOSTIC CENTRE**  
**NO:02 AMEER LAY OUT, VANDIMEDU,**  
**VILLUPURAM**  
**TAMILNADU605602**

Coordinator Contact Details:  
 Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
 Contact Number: 0416-2283102

*Pamela Christudoss*  
 Dr. Pamela Christudoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

**Data in CMC EQAS reports is confidential**  
**CMC EQAS does not sub contract any components**  
**\*\*\*\*\* End of Report \*\*\*\*\***