



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - JUNE 2023



PC-1074

Lab Name	Govt. UPHC, Vellakinar	Lab No	12059
Constituent Group	Chemistry III	Date of Result Entered	20/06/2023
PT Item	Lyophilized human serum based	Date of Report Published	04/07/2023

Sl. No.	Test Name	Method	Instrument	Result	Target	Delta	Delta 2	Delta 3	Delta 4	Delta 5	Delta 6
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	280 mg/dL	-0.45	0.69	
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.8 mg/dL	1.34	0.01	
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	116 mg/dL	0.45	0.24	

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. UPHC, Vellakinar
THE MEDICAL OFFICER, GOVT PRIMARY HEALTH CENTRE, VELLAKINAR
Coimbatore
TAMILNADU641029

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

மருத்துவ அலுவலர் (பொறுப்பு)
அரசு ஆரம்ப சுகாதார நிலையம்,
வெள்ளக்கிணர் - 641 029,
கோவை மாவட்டம்.