



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - JULY 2023



PC-1024

Lab Name: Govt. PHC, Irungalur
Constituent Group: Chemistry III
PT Item: Lyophilized human serum based

Lab No: 12724
Date of Result Entered: 20/07/2023
Date of Report Published: 07/08/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	103 mg/dL	-1.05	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	5.8 mg/dL	-0.09	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	112 mg/dL	-0.21	0.22
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :
Govt. PHC, Irungalur
Irungalur
TAMILNADU

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christodoss
Dr. Pamela Christodoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components
***** End of Report *****

MEDICAL OFFICER
Govt. Primary Health Centre
Irungalur, Trichy (Dt) 621 105.

[Signature]



CHRISTIAN MEDICAL COLLEGE
DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - AUGUST 2023



PC-1024

Lab Name Govt. PHC, Irungalur

Lab No 12724

Constituent Group Chemistry III

Date of Result Entered : 19/08/2023

PT Item Lyophilized human serum based

Date of Report Published : 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	179 mg/dL	0.38	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.9 mg/dL	1.18	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	116 mg/dL	0.47	0.22
SDI Range										
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						
Interpretation										

LAB ADDRESS :
Govt. PHC, Irungalur
Irungalur
TAMILNADU

Coordinator Contact Details:
Email: cllngqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Paula Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

MEDICAL OFFICER
Govt. Primary Health Centre
Irungalur, Trichy - 621 105



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - SEPTEMBER 2023



PC-1024

Lab Name

Govt. PHC, Irungalur

Lab No

12724

Constituent Group

Chemistry III

Date of Result Entered :

20/09/2023

PT item

Lyophilized human serum based

Date of Report Published :

05/10/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1870	310.86	4.82	14.98	302 mg/dL	-0.59	0.69
2	CREATININE III	Jafres Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.29	8.14	0.10	1.4 mg/dL	1.05	0.00
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1783	113.62	3.80	4.32	123 mg/dL	2.17	0.20
SDI Range Interpretation										
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :

Govt. PHC, Irungalur
Irungalur
TAMILNADU

Coordinator Contact Details:

Email: cilingqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

MEDICAL OFFICER
Govt. Primary Health Centre
Irungalur, Trichy (Dt)-621 105.