
















CALIBRATION / TESTING REPORT					
CUSTOMER DETAILS : GOVERNMENT PRIMARY HEALTH CENTRE SRPATTANAM, SIVAGANGAI.			CERTIFICATE NO		TICS/2023/SRP/SG/05
			EQUIPMENT NAME		PIPETTE
			MANUFACTURE		ERBA BIOHIT
			MODEL		10 µL FIXED
CALIBRATION DATE		20/10/2023	ASSET NO		SRP PH 05
CALIBRATION DUE		19/10/2024	DEPARTMENT		LAB
1. Have more dust in machine outside <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No 2. Any Physical damaged in machine <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No 3. Machine inside Cleaning (if required) <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No					
TEST EQUIPMENT		ID. NO	CERTIFICATE NUMBER		VALIDITY
ELECTRONIC SEMI MICRO BALANCE		NCS/WB01	TSC/23-24/5424-1		27-JUNE-2024
TEST REPORT					Temperature 25°C Humidity RH 55%
PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
VOLUME	µL	10	10.03	± 1	PASS
*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipments having traceability to National/International standards.					
<input checked="" type="checkbox"/> ACCEPTABLE FOR USE <input type="checkbox"/> CORRECTIVE MAINTENANCE REQUIRED <input type="checkbox"/> REMOVED FROM USE					
Calibrated By :			Authorized Sign With Seal		
					




CALIBRATION / TESTING REPORT					
CUSTOMER DETAILS : GOVERNMENT PRIMARY HEALTH CENTRE SRPATTANAM, SIVAGANGAI.		CERTIFICATE NO		TICS/2023/SRP/SG/03	
		EQUIPMENT NAME		MICROPIPETTE	
		MANUFACTURE		THERMO SCIENTIFIC	
		MODEL		FINE PIPETTE 100-1000 μ L	
CALIBRATION DATE	20/10/2023	ASSET NO		RW02595	
CALIBRATION DUE DATE	19/10/2024	INSTALLED DEPARTMENT		LAB	
<p>1. Have more dust in machine outside <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No</p> <p>2. Any Physical damaged in machine <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No</p> <p>3. Machine inside Cleaning (if required) <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No</p>					
TEST EQUIPMENT		ID. NO	CERTIFICATE NUMBER		VALIDITY
ELECTRONIC SEMI MICRO BALANCE		NCS/WB01	TSC/23-24/5424-1		27-JUNE-2024
TEST REPORT					Temperature 25°C Humidity RH 55%
PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
VOLUME	μ L	100	100.10	± 1	PASS
		1000	1000.08	± 1	PASS
*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipments having traceability to National/International standards.					
<input checked="" type="checkbox"/> ACCEPTABLE FOR USE <input type="checkbox"/> CORRECTIVE MAINTENANCE REQUIRED <input type="checkbox"/> REMOVED FROM USE					
Calibrated By :			Authorized Sign With Seal		
					
					

CALIBRATION / TESTING REPORT					
CUSTOMER DETAILS : GOVERNMENT PRIMARY HEALTH CENTRE SRPATTANAM, SIVAGANGAI.			CERTIFICATE NO		TICS/2023/SRP/SG/04
			EQUIPMENT NAME		MICROPIPETTE
			MANUFACTURE		THERMO SCIENTIFIC
			MODEL		FINE PIPETTE 10-100 µL
CALIBRATION DATE		20/10/2023	ASSET NO		RW01327
CALIBRATION DUE DATE		19/10/2024	INSTALLED DEPARTMENT		LAB
<p>1. Have more dust in machine outside <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No</p> <p>2. Any Physical damaged in machine <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No</p> <p>3. Machine inside Cleaning (if required) <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No</p>					
TEST EQUIPMENT		ID. NO	CERTIFICATE NUMBER		VALIDITY
ELECTRONIC SEMI MICRO BALANCE		NCS/WB01	TSC/23-24/5424-1		27-JUNE-2024
TEST REPORT					Temperature 25°C Humidity RH 55%
PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
VOLUME	µL	10	99.9	± 1	PASS
		100	99.99	± 1	PASS
*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipments having traceability to National/International standards.					
<input checked="" type="checkbox"/> ACCEPTABLE FOR USE <input type="checkbox"/> CORRECTIVE MAINTENANCE REQUIRED <input type="checkbox"/> REMOVED FROM USE					
Calibrated By :			Authorized Sign With Seal		
					
					

CALIBRATION / TESTING REPORT					
CUSTOMER DETAILS : GOVERNMENT PRIMARY HEALTH CENTRE SRPATTANAM, SIVAGANGAI.			CERTIFICATE NO		TICS/2023/SRP/SG/02
			EQUIPMENT NAME		CENTRIFUGE
			MANUFACTURE		BD INSTRUMENTATION
			MODEL		BDI 152
CALIBRATION DATE		20/10/2023	ASSET NO		SRPPH002
CALIBRATION DUE		19/10/2024	DEPARTMENT		LAB
<p>1. General maintenance required <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No</p> <p>2. Any Physical damage observed <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No</p> <p>3. Spare Replacement (if required) <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No</p>					
TEST EQUIPMENT		MANUFACTURE		MODEL	SERIAL NUMBER
DIGITAL TACHOMETER		KUSAM MECO		KM2234 BL	S1049875
TEST REPORT					Temperature 25°C Humidity RH 55%
PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
ROTATION PER MINUTE	RPM	SPEED 2	1889	± 5	PASS
		SPEED 5	2889		
*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipment having traceability to National/International standards.					
<input checked="" type="checkbox"/> ACCEPTABLE FOR USE <input type="checkbox"/> CORRECTIVE MAINTENANCE REQUIRED <input type="checkbox"/> REMOVED FROM USE					
Calibrated By :			Authorized Sign With Seal		
					
					

CALIBRATION / TESTING REPORT				
CUSTOMER DETAILS : GOVERNMENT PRIMARY HEALTH CENTRE SRPATTANAM, SIVAGANGAI.		CERTIFICATE NO	TICS/2023/SRP/SG/01	
		EQUIPMENT NAME	SEMI ANALYZER	
		MANUFACTURE	ADAM AG	
		MODEL	PRIETEST TOUCH	
CALIBRATION DATE	20/10/2023	SERIAL NO	AT1970117SMD	
CALIBRATION DUE DATE	19/10/2024	INSTALLED DEPARTMENT	LAB	
1. General maintenance required <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No 2. Any Physical damage observed <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No 3. Spare Replacement (if required) <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No				
TEST EQUIPMENT	MANUFACTURE	MODEL	SERIAL NUMBER	
ELECTRICAL SAFTY ANALYZER	DATREND	V Pad-mini	VPM 18040011	
PARAMETER	UNITS	VALUE	LIMITS	STATUS
Main Voltage L1-L2	V	227.8	240±10	PASS
Main Voltage L1-G(PE)	V	226.1	240±10	PASS
Main Voltage L2-G(PE)	V	1.2	<5	PASS
Load Current	A	0.402	As per unit	PASS
Protective Earth	Ω	0	<0.3	PASS
Point-to-Point Resistance	Ω	0	<0.5	PASS
Equipment Leakage	μA	0.11	Class I - < 500 Class II - < 100	PASS
Applied Part Leakage	μA	0	Type BF - < 5mA Type CF- <50μA*	PASS
Note : Test Standard: IEC 62353 Defibrillator paddles 100 μA, per IEC 60601-2-4				
*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipment having traceability to National/International standards.				
<input checked="" type="checkbox"/> ACCEPTABLE FOR USE <input type="checkbox"/> CORRECTIVE MAINTENANCE REQUIRED <input type="checkbox"/> REMOVED FROM USE				
Calibrated By : Authorized Sign With Seal				
				
				

CALIBRATION / TESTING REPORT				
CUSTOMER DETAILS : GOVERNMENT PRIMARY HEALTH CENTRE SRPATTANAM, SIVAGANGAL.		CERTIFICATE NO	TICS/2023/SRP/SG/06	
		EQUIPMENT NAME	MICROSCOPE	
		MANUFACTURE	NA	
		MODEL	N101C	
CALIBRATION DATE	20/10/2023	SERIAL NO	01282	
CALIBRATION DUE DATE	19/10/2024	INSTALLED DEPARTMENT	LAB	
1. General maintenance required <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No 2. Any Physical damage observed <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No 3. Spare Replacement (if required) <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No				
TEST EQUIPMENT	MANUFACTURE	MODEL	SERIAL NUMBER	
ELECTRICAL SAFTY ANALYZER	DATREND	V Pad-mini	VPM 18040011	
PARAMETER	UNITS	VALUE	LIMITS	STATUS
Main Voltage L1-L2	V	226.8	240±10	PASS
Main Voltage L1-G(PE)	V	226.1	240±10	PASS
Main Voltage L2-G(PE)	V	1.2	<5	PASS
Load Current	A	0.122	As per unit	PASS
Protective Earth	Ω	0	<0.3	PASS
Point-to-Point Resistance	Ω	0	<0.5	PASS
Equipment Leakage	μA	0	Class I - < 500 Class II - < 100	PASS
Applied Part Leakage	μA	0	Type BF - < 5mA Type CF- <50μA *	PASS
Note : Test Standard: IEC 62353 Defibrillator paddles 100 μA, per IEC 60601-2-4				
*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipment having traceability to National/International standards.				
<input checked="" type="checkbox"/> ACCEPTABLE FOR USE <input type="checkbox"/> CORRECTIVE MAINTENANCE REQUIRED <input type="checkbox"/> REMOVED FROM USE				
Calibrated By : 			Authorized Sign With Seal	
