



Certificate No : PHC/MP/027

CERTIFICATE OF CALIBRATION

Customer Details	Details of DUC	
URBAN PRIMARY HEALTH CENTRE, KUZHITHURAI, VILAVANCODE, KANNIYAKUMARI, TAMILNADU- 629163	Description	MICROPIPETTE
	Manufactured by	NIL
	Model	5-50 μ l
	Serial No.	NIL
	Asset ID	NIL
	Location	LABORATORY

Environmental Conditions		Calibration Details	
Operating Temperature	27 $^{\circ}$ C	Date of Receipt	30/10/2023
		Date of Calibration	30/10/2023
Humidity	56%	Calibration Due Date	29/10/2024
		Date of Issue	31/10/2023
		Cond. On Receipt	Satisfactory

Details of Calibration Standards

Name	Manufacturer	Serial No.	Valid up to	Traceability
Electronic Semi Micro Balance	AND, GR-202	14251938	27-DEC-2023	Equations

Performance Analysis

SI.No	Device Reading(μ l)	Mean Value(μ l)	Deviation
1.	10	10.1	0
2.	5	5.2	0.1
3.	1	0.9	0.1



Certificate No : PHC/MP/027

Remarks:

- *DUC is defined as device under calibration.*
- *Measured value given was the average of 10 measurements.*
- *Master equipment used for calibration were calibrated from accredited laboratory & traceable to National & International standards.*
- *The calibration results reported are valid at the time of and under the stated conditions of the measurements.*
- *The calibration certificate shall not be reproduced except in full without written approval of Biomedical Techniques.*
- *The above results relate only to the item calibrated.*
- *The calibration certificate shall not be utilized for any legal purpose produced in the court of law for which no responsibility will be associated to Biomedical Techniques.*

Done By

A handwritten signature in blue ink, appearing to read "Amal Nath K J", written over a horizontal line.



AMALNATH K J
Calibration Engineer

Authorized By

A handwritten signature in blue ink, appearing to read "Remya CR", written over a horizontal line.

REMYA CR
Technical Manager

***End of certificate ***