



Lab Name Govt. Block PHC, Malliakarai  
Constituent Group Chemistry III  
PT item Lyophilized human serum based

Lab No 7488  
Date of Result Entered : 20/04/2023  
Date of Report Published : 06/05/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1832	119.45	6.04	7.21	123 mg/dL	0.49	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1820	2.16	8.09	0.18	2.1 mg/dL	-0.34	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1750	111.61	5.19	5.79	121 mg/dL	1.62	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

## LAB ADDRESS :

Govt. Block PHC, Malliakarai  
THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, SALEM DISTRICT.  
Malliakarai  
TAMILNADU636107

Coordinator Contact Details:  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact Number: 0416-2283102

*Pamela Christudoss*

Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential

[https://home.cmcvellore.ac.in/clinqc/ViewMonthlySummary/printResult\\_new.aspx?resDetId=264&noSamples=1&month=4&year=2023&groupId=9&yearGroup=110](https://home.cmcvellore.ac.in/clinqc/ViewMonthlySummary/printResult_new.aspx?resDetId=264&noSamples=1&month=4&year=2023&groupId=9&yearGroup=110)



Lab Name: Govt. Block PHC, Malliakarai

Lab No: 7488

Constituent Group: Chemistry III

Date of Result Entered: 20/05/2023

PT Item: Lyophilized human serum based

Date of Report Published: 05/06/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1834	180.88	5.19	9.39	171 mg/dL	-1.05	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate	Robonik ( Semi automatic / Fully Automatic )	1854	1.76	7.48	0.13	1.5 mg/dL	-1.97	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1798	114.60	5.08	5.83	119 mg/dL	0.76	0.27

SDI Range		Interpretation
Within -1.00 to +1.00		Excellent.
Within ±1.01 to ±2.00		Good.
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.
Beyond ±3.0		Unacceptable performance. Action Signal.

**LAB ADDRESS :**  
 Govt. Block PHC, Malliakarai  
 THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, SALEM DISTRICT.  
 Malliakarai  
 TAMILNADU636107

**Coordinator Contact Details:**  
 Email: [clnqc@cmcvellore.ac.in](mailto:clnqc@cmcvellore.ac.in)  
 Contact Number: 0416-2283102

*Pamela Christudoss*  
 Dr. Pamela Christudoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential



MONTHLY SUMMARY REPORT - JUNE 2023

PC-1024

Lab Name: Govt. Block PHC, Malliakarai  
 Lab No: 7488

Constituent Group: Chemistry III  
 Date of Result Entered: 20/06/2023

PT item: Lyophilized human serum based  
 Date of Report Published: 04/07/2023

SI.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	283 mg/dL	-0.24	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.7 mg/dL	0.76	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	115 mg/dL	0.25	0.24
SDI Range Interpretation										
Within -1.00 to +1.00				Excellent						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :

Govt. Block PHC, Malliakarai  
 THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, SALEM DISTRICT.  
 Malliakarai  
 TAMILNADU636107

Coordinator Contact Details:  
 Email:clinqc@cmcvellore.ac.in  
 Contact Number: 0416-2283102

*Pamela Christodoss*  
 Dr. Pamela Christodoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
 Data In CMC EQAS reports is confidential

home.cmcvellore.ac.in/clinqc/ViewMonthlySummary/printResult\_new.aspx?resDetId=266&nosamples=1&month=6&year=2023&groupid=9&yearGroup=110

*ad.*  
*1.01.23*  
*Microbiology*



**CHRISTIAN MEDICAL COLLEGE**  
DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - JULY 2023



PC-1024

Lab Name Govt. Block PHC, Mallakaral Lab No 7488  
 Constituent Group Chemistry III Date of Result Entered : 20/07/2023  
 PT Item Lyophilized human serum based Date of Report Published : 07/08/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants			Your Value	SDI	U
						CV	SD				
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	117 mg/dL	1.26	0.28	
2	CREATININE III	Jaffes Kinetic - Alkaline Pictate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	5.7 mg/dL	-0.37	0.02	
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	109 mg/dL	-0.86	0.22	

SDI Range		Interpretation
Within -1.00 to +1.00		Excellent.
Within $\pm 1.01$ to $\pm 2.00$		Good.
Within $\pm 2.01$ to $\pm 2.99$		Accept with caution. Warning Signal.
Beyond $\pm 3.0$		Unacceptable performance. Action Signal.

**LAB ADDRESS :**

Govt. Block PHC, Mallakaral  
 THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, SALEM DISTRICT.  
 Mallakaral  
 TAMILNADU636107

*Paula Christodoss*

Coordinator Contact Details:  
 Email:clnqcc@cmcvellore.ac.in  
 Contact Number: 0416-2283102

Dr. Pamela Christudoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
 Data in CMC EQAS reports is confidential  
 CMC EQAS does not sub contract any components  
 \*\*\*\*\* End of Report \*\*\*\*\*

## VIEW LAB MONTHLY SUMMARY

<b>Lab Name</b>	Govt. Block PHC, Mallakara				<b>Details About Robust Analysis</b>						
<b>Lab No</b>	7488				<b>Detail About Monthly Summary</b>						
<b>Month</b>	August				<b>Detail about SDI</b>						
<b>Year</b>	2023										
<b>Constituent Group</b>	Chemistry III										
				Click on the analyte to view Graphical Data				<a href="#">All Analyser Result</a>		<a href="#">Print</a>	
								Date of Result Entered : 19/08/2023			
								Date of Report Published : 01/09/2023			
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U	
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1841	175.58	5.13	9.00	187 mg/dL	1.27	0.42	
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.76	6.77	0.12	1.7 mg/dL	-0.50	0.01	
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1764	113.80	4.07	4.64	117 mg/dL	0.69	0.22	
SDI Range				Interpretation							
Within -1.00 to +1.00				Excellent.							
Within ±1.01 to ±2.00				Good.							
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.							
Beyond ±3.0				Unacceptable performance. Action Signal.							

## Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>

## VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. Block PHC, Mallakeral	Details About Robust Analysis								
Lab No	7488	Detail About Monthly Summary								
Month	September	Detail about SDI								
Year	2023									
Constituent Group	Chemistry III									
Click on the analyte to view Graphical Data		<a href="#">All Analyser Result</a>	<a href="#">Print</a>							
Date of Result Entered : 20/09/2023										
Date of Report Published : 05/10/2023										
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1870	310.86	4.82	14.98	329 mg/dL	1.21	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.29	8.14	0.10	1.2 mg/dL	-0.88	0.00
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1783	113.62	3.80	4.32	118 mg/dL	1.01	0.20
SDI Range		Interpretation								
Within -1.00 to +1.00		Excellent.								
Within $\pm 1.01$ to $\pm 2.00$		Good.								
Within $\pm 2.01$ to $\pm 2.99$		Accept with caution. Warning Signal.								
Beyond $\pm 3.0$		Unacceptable performance. Action Signal.								

## Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>