

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. UPHC, Seetharam Nagar	Details About Robust Analysis
Lab No	12345	Detail About Monthly Summary
Month	April	Detail about SDI
Year	2023	
Constituent Group	Chemistry III	

[Click on the analyte to view Graphical Data](#)

All Analyser Result
Print

Date of Result Entered : 20/04/2023

Date of Report Published : 06/05/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	Participants SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	121 mg/dL	0.21	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.4 mg/dL	1.37	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	111 mg/dL	-0.11	0.28
SDI Range		Interpretation								
Within -1.00 to +1.00		Excellent.								
Within ±1.01 to ±2.00		Good.								
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. UPHC, Seetharam Nagar	Details About Robust Analysis
Lab No	12345	Detail About Monthly Summary
Month	May	Detail about SDI
Year	2023	
Constituent Group	Chemistry III	

[Click on the analyte to view Graphical Data](#)

All Analyser Result
Print

Date of Result Entered : 18/05/2023

Date of Report Published : 05/06/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1834	180.88	5.19	9.39	204 mg/dL	2.46	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1854	1.76	7.48	0.13	1.9 mg/dL	1.06	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.83	199 mg/dL	14.48	0.27
SDI Range			Interpretation							
Within -1.00 to +1.00			Excellent.							
Within ±1.01 to ±2.00			Good.							
Within ±2.01 to ±2.99			Accept with caution. Warning Signal.							
Beyond ±3.0			Unacceptable performance. Action Signal.							

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. UPHC, Seetharam Nagar	Details About Robust Analysis
Lab No	12345	Detail About Monthly Summary
Month	June	Detail about SDI
Year	2023	
Constituent Group	Chemistry III	

[Click on the analyte to view Graphical Data](#)

Date of Result Entered : 20/06/2023

Date of Report Published : 04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	269 mg/dL	-1.19	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.5 mg/dL	-0.41	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	119 mg/dL	1.04	0.24
SDI Range			Interpretation							
Within -1.00 to +1.00			Excellent.							
Within ±1.01 to ±2.00			Good.							
Within ±2.01 to ±2.99			Accept with caution. Warning Signal.							
Beyond ±3.0			Unacceptable performance. Action Signal.							

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. UPHC, Seetharam Nagar	Details About Robust Analysis
Lab No	12345	Detail About Monthly Summary
Month	July	Detail about SDI
Year	2023	
Constituent Group	Chemistry III	<input type="text"/> <input type="text"/>
Click on the analyte to view Graphical Data		All Analyser Result Print
Date of Result Entered : 19/07/2023		
Date of Report Published : 07/08/2023		

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	107 mg/dL	-0.39	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	6.2 mg/dL	1.06	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	112 mg/dL	-0.21	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. UPHC, Seetharam Nagar	Details About Robust Analysis
Lab No	12345	Detail About Monthly Summary
Month	August	Detail about SDI
Year	2023	
Constituent Group	Chemistry III	

[Click on the analyte to view Graphical Data](#)

All Analyser Result
Print

Date of Result Entered : 17/08/2023

Date of Report Published : 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	173 mg/dL	-0.29	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.6 mg/dL	-1.34	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	124 mg/dL	2.20	0.22
SDI Range		Interpretation								
Within -1.00 to +1.00		Excellent.								
Within ±1.01 to ±2.00		Good.								
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. UPHC, Seetharam Nagar	Details About Robust Analysis
Lab No	12345	Detail About Monthly Summary
Month	September	Detail about SDI
Year	2023	
Constituent Group	Chemistry III	

[Click on the analyte to view Graphical Data](#)

All Analyser Result
Print

Date of Result Entered : 19/09/2023

Date of Report Published : 05/10/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1870	310.86	4.82	14.98	310 mg/dL	-0.06	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.29	8.14	0.10	3.1 mg/dL	18.1	0.00
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1783	113.62	3.80	4.32	117 mg/dL	0.78	0.20
SDI Range			Interpretation							
Within -1.00 to +1.00			Excellent.							
Within ±1.01 to ±2.00			Good.							
Within ±2.01 to ±2.99			Accept with caution. Warning Signal.							
Beyond ±3.0			Unacceptable performance. Action Signal.							

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out