



CHRISTIAN MEDICAL COLLEGE
 DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMG EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - JUNE 2023



Lab Name: Govt. Block PHC, Uppilipapuram
 Constituent Group: Chemistry III
 PT Item: Lyophilized human serum based

Lab No: 7246
 Date of Result Entered: 20/06/2023
 Date of Report Published: 04/07/2023

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	SD	Your Value	SD	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	278 mg/dL	-0.58	0.69
2	UREA III	Others (Urease Berthelot / Ned Dye) III	Any Analyser (Automation / Semi Automation)	50	33.99	14.82	5.04	56 mg/dL	4.37	1.42
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.8 mg/dL	1.34	0.01
4	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	113 mg/dL	-0.14	0.24
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution, Warning Signal.						
Beyond ±3.0				Unacceptable performance, Action Signal.						

LAB ADDRESS :
 Govt. Block PHC, Uppilipapuram
 THE BLOCK MEDICAL OFFICER, TRICHY DISTRICT
 Uppilipapuram
 TAMILNADU621011

Coordinator Contact Details:
 Email:clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Tanya Christodoss
 Dr. Pamela Christodoss
 CMG EQAS Coordinator
 Christian Medical College, Vellore