



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JULY 2023



Lab Name: Govt. Block PHC, Uppiliapuram

Constituent Group: Chemistry III

PT item: Lyophilized human serum based

Lab No: 7246

Date of Result Entered: 18/07/2023

Date of Report Published: 07/08/2023

Sl No	Analyte	Method / Principle Name	Analyse Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	CV
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	109 mg/dL	-0.06	0.28
2	UREA III	Others (Urease Berthelot / Ned Dye) III	Any Analyser (Automation / Semi Automatic)	47	52.80	7.28	3.85	74 mg/dL	5.51	1.12
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	6.08 mg/dL	0.71	0.02
4	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	102.3 mg/dL	-2.30	0.22
Interpretation										
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :
 Govt. Block PHC, Uppiliapuram
 THE BLOCK MEDICAL OFFICER, TRICHY DISTRICT
 Uppiliapuram
 TAMILNADU621011

Coordinator Contact Details:
 Email: cllinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Paola Christodou
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore