



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - AUGUST 2023

Lab Name Govt. Block PHC, Uppiliapuram

Constituent Group Chemistry III

PT item Lyophilized human serum based

Lab No

Date of Result Entered :

Date of Report Published :

SI No Analyte Method / Principle Name

Analyzer Name

No of Participants

DV

Participants CV

SD

Unit

SI No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Unit
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1841	175.58	5.13	9.00	188 mg/dl
2	UREA III	Others ( Urease Berthelot / Ned Dye ) III	Any Analyser (Automatic / Semi Automatic )	59	28.92	13.99	4.05	36 mg/dl
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.76	6.77	0.12	1.5 mg/dl
4	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1764	113.80	4.07	4.64	115 mg/dl
SDI Range				Interpretation				
Within -1.00 to +1.00				Excellent.				
Within ±1.01 to ±2.00				Good.				
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.				
Beyond ±3.0				Unacceptable performance. Action Signal.				

## LAB ADDRESS :

Govt. Block PHC, Uppiliapuram  
THE BLOCK MEDICAL OFFICER, TRICHY DISTRICT  
Uppiliapuram  
TAMILNADU621011

*Pavula Christudasa*

Dr. Pavula Christudasa