

				CALIE	3RA	TION	CERT	FICATE		
CERTIFICATE NO: SBS/CL/23/10363			3/10363	/10363		MECHANICAL		Pag	e No:1 of 1	
Issue Date				18-08-2023						
SRF No & Date				SRF/23/00320-0002 & 17-08-2023						
Receipt Date					17-08-2023					
Calibration Date					17-08-2023					
Calibra	Calibration Due					16-08-2024				
Custo	mer Name 8	& Addre	ess							
GOV⊞	RYWYJAEHWIT 197	NYMAMO	VHEAHEAC	ENTRENTRE,						
REDAK	PAELINANTHAJW	651RUF	PAUPURTO	ISTRIBOTRICT.						
				Deta	ails of D	evice Un	der Calibrati	ion (DUC)		
Descri	scription : CENTRIFUGE			Make & Model		;	BD INSTRUMENTATION & BD-153			
Range	lange : 5000 RPM			Sr. No		1	76659			
Resolution : 50 RPM			Identification No		:	NA				
DUC C	DUC Condition : Satisfactory			Location			:	LABORATORY		
			Envi	ironmental C	ondition	ns & Star	ndard Operat	ting Procedur	e Details	
Environmental Details Temperature:				: 24.3 °C Re		Relative Humidity		54 % Rh		
Calibration Procedure No SBS/CP/ML/0				04 Calibration		Calibration d	done at ONSITE			
1					Refer	ence Sta	ndards Deta	ils		
S.No	lo Description			Make/ SI No:		Certificate No		Validity		
1	Digital Tachometer			Lutron & Q638380		JRPM-CCTR-A&S-2022-0036		03-10-2023		

CALIBRATION RESULTS

S.No	DEVICE UNDER CALIBRATION READINGS	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY (±)	
	RPM	· RPM	RPM	%	
1	1000	999.5	0.5	4.2	
2	2000	1999.7	0.3	4.2	
3	3000	2999.5	0.5	4.2	

REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By	STILL CHO CO	Authorised Signator	Authorised Signatory		
(Calibration Engineer) M.RAGUL	SNOW SUNDAN A	Quality Manager Chief Execu	tive		

SUNSHINE BIOMEDICAL SOLUTIONS