



CALIBRATION / TESTING REPORT			
CUSTOMER DETAILS :		CERTIFICATE NO	TICS/2023/OSV/SG/02
		EQUIPMENT NAME	CENTRIFUGE
		MANUFACTURE	NA
		MODEL	NA
CALIBRATION DATE	20/10/2023	ASSET NO	OSVPH02
CALIBRATION DUE	19/10/2024	DEPARTMENT	LAB

1. General maintenance required Yes No

2. Any Physical damage observed Yes No

3. Spare Replacement (if required) Yes No

TEST EQUIPMENT	MANUFACTURE	MODEL	SERIAL NUMBER
DIGITAL TACHOMETER	KUSAM MECO	KM2234 BL	S1049875

TEST REPORT Temperature 25°C
Humidity RH 55%

PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
ROTATION PER MINUTE	RPM	SPEED 2	1690	± 5	PASS
		SPEED 4	2317		

*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipment having traceability to National/International standards.

ACCEPTABLE FOR USE CORRECTIVE MAINTENANCE REQUIRED REMOVED FROM USE

Calibrated By:  Authorized Sign With Seal 

