

CALIBRATION / TESTING REPORT					
CUSTOMER DETAILS : GOVERNMENT PRIMARY HEALTH CENTRE O SIRUVAYAL, SIVAGANGAI.		CERTIFICATE NO		TICS/2023/OSV/SG/05	
		EQUIPMENT NAME		MICROPIPETTE	
CALIBRATION DATE		20/10/2023		ASSET NO	
		19/10/2024		OSV PH 05	
CALIBRATION DUE DATE		19/10/2024		INSTALLED DEPARTMENT	
				LAB	
<p>1. Have more dust in machine outside <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No</p> <p>2. Any Physical damaged in machine <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No</p> <p>3. Machine inside Cleaning (if required) <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No</p>					
TEST EQUIPMENT		ID. NO		CERTIFICATE NUMBER	
ELECTRONIC SEMI MICRO BALANCE		NCS/WB01		TSC/23-24/5424-1	
				27-JUNE-2024	
TEST REPORT					Temperature 25°C Humidity RH 55%
PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
VOLUME	μL	10	99.98	± 1	PASS
		100	100.09	± 1	PASS
*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipments having traceability to National/International standards.					
<input checked="" type="checkbox"/> ACCEPTABLE FOR USE <input type="checkbox"/> CORRECTIVE MAINTENANCE REQUIRED <input type="checkbox"/> REMOVED FROM USE					
Calibrated By : 					
					