



CALIBRATION / TESTING REPORT					
CUSTOMER DETAILS :			CERTIFICATE NO	TICS/2023/OSV/SG/04	
GOVERNMENT PRIMARY HEALTH CENTRE O SIRUVAYAL, SIVAGANGAL.			EQUIPMENT NAME	MICROPIPETTE	
			MANUFACTURE	MICROLUX	
			MODEL	100-1000 µL	
CALIBRATION DATE	20/10/2023		ASSET NO	OSVPH 04	
CALIBRATION DUE DATE	19/10/2024		INSTALLED DEPARTMENT	LAB	
<p>1. Have more dust in machine outside <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No</p> <p>2. Any Physical damaged in machine <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No</p> <p>3. Machine inside Cleaning (if required) <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No</p>					
TEST EQUIPMENT	ID. NO	CERTIFICATE NUMBER	VALIDITY		
ELECTRONIC SEMI MICRO BALANCE	NCS/WB01	TSC/23-24/5424-1	27-JUNE-2024		
TEST REPORT					Temperature 25°C Humidity RH 55%
PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
VOLUME	µL	100	100.04	± 1	PASS
		1000	1000.03	± 1	PASS
*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipments having traceability to National/International standards.					
<input checked="" type="checkbox"/> ACCEPTABLE FOR USE <input type="checkbox"/> CORRECTIVE MAINTENANCE REQUIRED <input type="checkbox"/> REMOVED FROM USE					
Calibrated By :			Authorized Sign With Seal		

