

Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone : 0452 - 2561691 Mobile : +91 94434 74610



GSTIN 33AAHCA8053H1ZR • DL No. MDU/4205/208,MDU/3971/21B • CIN : U33110TN2009PTC072140

CALIBRATION CERTIFICATE



Customer Name & Address	M/S. Urban Primary Health Centre, UPHC - 25 Kaliyaman Kovil Street, Chinthamani, Madurai - 625 009.	Certificate No:	AMS/HOSP/2889306
		Issuing Date:	02/03/2021
		Page No:	1 of 1

Details of Device under Calibration (DUC):

DUC:	Ice Lined Refrigerator	Id. No:	CH/UPHC/LR/ILR/001
Make:	Vest Frost Solutions	Department:	Labour Room
Model:	VLS - 300	Calibrated On:	01/03/2021
Sl No:	20172001429	Calibration Due Suggested:	28/02/2022
Type:	Indirect Cool	DUC Condition on Receipt:	OK
Range & Least Count:	Minus 10 - 8°C/ LC 1°C		

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	RTD Thermometer	Center / RTD 375	PC/01/21/ATL/0107-01	22/01/2022
2	Thermo Hygrometer	Beetech/088	PC/01/21/ATL/0194-01	22/01/2022

Room Temperature: 25°C±2 Room Humidity: 30-60% (At the time of Calibration)

Test Results:

Sl.No.	Place	DUC Reading (°C)	Standard Reading (°C)	Correction (°C)
5	Chamber I	3.3	2.98	-0.32
6		3.5	2.66	-0.84
7	Chamber II	3.4	2.83	-0.57
8		3.6	3.38	-0.22
9	Chamber III	3.2	3.12	-0.08
10		3.5	3.41	-0.09

Allowable Tolerance: User Defined.

Note

1. Calibration Points were selected as per customer request.
2. Each reading were taken by Repeatability Test (i.e) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:
(Calibration Engineer)

B. Dey



Approved by:
(Technical Manager)

B. Dey

