

## CENTRE for HIV TESTING State Reference Laboratory for HIV Department of Microbiology Lady Hardinge Medical College & Associated Hospitals New Delhi

Ref. No. MTON 2022 930 | CAT-LAMC 55

Date: 31/10/22

## **EQAS REPORT (SRL to ICTCs/PPTCTs)**

Round: I

Name of Proficiency Testing provider: SRL, CHT-LHMC Date of panel distribution: 21/10/22

Name of the participating laboratory: Giridhar Lal Maternity Hospital (PPTCT)

Date of result received: 26/10/22

Sample code	Standard result	Results obtained by participating laboratory	% matching
A3	Positive for HIV-1 abs	Positive for HIV-1	100
A4	Negative for HIV abs	Negative	100
A1	Positive for HIV-1 abs	Positive for HIV-1	100
A2	Negative for HIV abs	Negative	100

Remarks: Performance is satisfactory.

AT. V. S. Randhawa

Name of Laboratory In-Charge:

Signature of Laboratory In Charge

\*\*\* End of Report\*\*\*



## CENTRE for HIV TESTING State Reference Laboratory for HIV Department of Microbiology Lady Hardinge Medical College & Associated Hospitals

Ref. No. Mrcsv 2022/930/ CHT- LHMC/ 55.

Date: 31/10/22

## **EQAS REPORT (SRL to ICTCs/PPTCTs)**

Round: I

Name of Proficiency Testing provider: SRL, CHT-LHMC Date of panel distribution: 21/10/22

Name of the participating laboratory: Giridhar Lal Maternity Hospital (ICTC)

Date of result received: 26/10/22

Sample code	Standard result	Results obtained by participating laboratory	% matching
A2	Positive for HIV-1 abs	Positive for HIV-1	100
А3	Negative for HIV abs	Negative	100
A4	Positive for HIV-1 abs	Positive for HIV-1	100
A1	Negative for HIV abs	Negative	100

Remarks: Performance is satisfactory.

Ar. v. s. Randhawa

Name of Laboratory In-Charge:

Signature of Laboratoevin Charge aboratory (CHT)

State Refor And State Reforment of Micro 110001

Department of New Delhi-110001

\*\*\* End of Report\*\*\*