

CHRISTIAN MEDICAL COLLEGE

CMC EXTERNAL QUALITY ASSURANCE SCHEME

2022 EQAS YEARLY SUMMARY

The yearly summary is based on the overall mean SDI of every analyte. The groups are made depending upon the number of results received for every individual analyte. The rank is mentioned for every analyte within the group. This modification in the yearly summary is done with the guidelines provided by NABL during the assessment for accreditation.

Group I : Laboratories who reported 10 to 12 months during 2022

Group II : Laboratories who reported 6 to 9 months during 2022

Group III : Laboratories who reported less than 6 months during 2022

Sl.No	Parameter	Rank Group	No of Labs	OMSDI	Grade
1	GLUCOSE III	Group I	1941	0.90	A
2	CREATININE III	Group I	1927	0.99	A
3	CHOLESTROL III	Group I	1923	1.47	B

Only first 2 Groups are Graded

OMSDI Range	Grade	Remarks
0.00 to 1.00	A	Excellent
1.01 to 2.00	B	Good
2.01 to 2.99	C	Satisfactory
Beyond \pm 3.00	D	Unacceptable

Your Lab No 10534

Constituent Group Chemistry III

Address

Govt. PHC, Manampathy ,
CLINICAL LABORATORY, KANCHIPURAM(DIST) ,
Manampathy ,
TAMILNADU ,603105 .

Print

Links

Modify Address

Modify Methods

Submit Lab Results

View Lab Monthly Summary

Yearly Summary

Change Password

Home

Sign-out

J. M. J. J. J.
ASSISTANT SURGEON
GOVT. PRIMARY HEALTH CENTRE
MANAMPATHY - 603 105.

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Manampathy
 Lab No 10534
 Month March
 Year 2023
 Constituent Group Chemistry III

Details About Robust Analysis
 Detail About Monthly Summary
 Detail about SDI

[All Analyser Result](#) [Print](#)

Date of Result Entered : 19/03/2023

Date of Report Published : 04/04/2023

[Click on the analyte to view Graphical Data](#)

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1878	91.48	7.46	6.82	87 mg/dL	-0.66	0.31
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1851	1.33	9.08	0.12	1.3 mg/dL	-0.25	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1761	110.93	5.23	5.80	111 mg/dL	0.01	0.28
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

[Signature]
 ASSISTANT DIRECTOR
 GOVT. PRIMARY HEALTH CENTRE
 MANAMPATHY - 603 105.

Links

[Modify Address](#)

[Modify Methods](#)

[Submit Lab Results](#)

[View Lab Monthly Summary](#)

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Manampathy

[Details About Robust Analysis](#)

Lab No 10534

[Detail About Monthly Summary](#)

Month April

[Detail about SDI](#)

Year 2023

Constituent Group Chemistry III

[Click on the analyte to view Graphical Data](#)

Date of Result Entered : 15/04/2023

Date of Report Published : 06/05/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	114 mg/dL	-0.76	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	1.9 mg/dL	-1.49	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	119 mg/dL	1.28	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

S. N. J. R.
 ASSISTANT SURGEON
 GOVT. PRIMARY HEALTH CENTRE
 MANAMPATHY - 603 105.

[Links](#)

[Modify Address](#)

[Modify Methods](#)

[Submit Lab Results](#)

[View Lab Monthly Summary](#)

VIEW LAB MONTHLY SUMMARY

Lab Name Govt PHC, Manampatty

Lab No 10534

Month May

Year 2023

Constituent Group Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)


[All Analyser Result](#)

[Print](#)

Date of Result Entered : 18/05/2023

Date of Report Published : 05/06/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1834	180.88	5.19	9.39	186 mg/dL	0.55	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1854	1.76	7.48	0.13	1.7 mg/dL	-0.45	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.83	111 mg/dL	-0.62	0.27
SDI Range Interpretation										
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						


 ASSISTANT SURGEON
 GOVT. PRIMARY HEALTH CENTRE
 MANAMPATTY - 602 404

[Links](#)

[Modify Address](#)

[Modify Methods](#)

[Submit Lab Results](#)

[View Lab Monthly Summary](#)

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Manampathy
 Lab No 10534
 Month June
 Year 2023
 Constituent Group Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 19/06/2023

Date of Report Published : 04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	308 mg/dL	1.44	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	26 mg/dL	0.18	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	115 mg/dL	0.25	0.24
SDI Range Interpretation										
				Within -1.00 to +1.00	Excellent.					
				Within ±1.01 to ±2.00	Good.					
				Within ±2.01 to ±2.99	Accept with caution. Warning Signal.					
				Beyond ±3.0	Unacceptable performance. Action Signal.					

S. N. J. Pr.
 ASSISTANT SURGEON
 GOVT. PRIMARY HEALTH CENTRE
 MANAMPATHY - 688 105.

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Manampathy

Lab No 10534

Month July

Year 2023

Constituent Group Chemistry III

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 16/07/2023

Date of Report Published : 07/08/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	99 mg/dL	-1.71	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	5.7 mg/dL	-0.37	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	119 mg/dL	1.29	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links	Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary		Change Password	Home	Sign-out

J. M. J. J. J.
 ASSISTANT SURGEON
 GOVT. PRIMARY HEALTH CENTRE
 MANAMPATHY - 603 105.

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Manampathy
 Lab No 10534
 Month August
 Year 2023
 Constituent Group Chemistry III

Details About Robust Analysis
 Detail About Monthly Summary
 Detail about SDI

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 19/08/2023

Date of Report Published : 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	189 mg/dL	1.49	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	0.77	0.12	1.6 mg/dL	-1.34	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	111 mg/dL	-0.60	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

S. N. J. P.
 ASSISTANT SURGEON
 GOVT. PRIMARY HEALTH CENTRE
 MANAMPATHY - 603 105.