## **AMS Healthcare**

# (Erstwhile: Ace Medical Solutions (India) Pvt. Ltd.)

An ISO 9001-2015 Certified Company





GSTIN. 33AOSPA9153Q1ZA • DL No. TN-12-20B -00398, TN-12-21B-00398

### **CALIBRATION CERTIFICATE**

Customer	Customer M/S. Bharathi Laboratory,		AMS/HSP/289/5191
	#69/A, Swami Sannathi Street,	Issuing Date:	30/10/2023
Address:	Tirunelveli Town – 6.	Page No:	1 of 1

**Details of Device under Calibration (DUC):** 

DUC:	Micro Pipette	Id. No:	BH/LAB/FMP/001
Make:	Handypette	Department:	Laboratory
Model:	·	Calibrated On:	28/10/2023
Sl. No:	223625439	Calibration Due	
Type:	Fixed	Suggested:	27/10/2024
Range &	10μl	DUC Condition	
Least Count:	•	on Receipt:	OK

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Electronic Weighing Scale	EWC/SF400C	PC/01/23/ATL/0512-02	31/01/2024
2	Thermo Hygrometer	Beetech/088	PC/01/23/ATL/0210-02	31/01/2024

Room Temperature: 25°C±2

Room Humidity: 30-60%

(At the time of Calibration)

#### **Test Results:**

SL.	DUC Reading	Standard Reading (µl)		Correction
No.	(μl)	Trial 1	Trial 2	(µl)
1	10	10.00	10.00	0

Allowable Tolerance: User Defined

#### Note:

- 1. Calibration Points were selected as per customer request.
- 2. Each reading were taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
- 3. Results Reported are valid at the time of and under the stated conditions of measurement

Calibrated By:

(Calibration Engineer)

amshealthcwemdu@gmail.com

Approved by: (Technical Manager)

94434 74610, 93455 56148

aceagm@gmail.com