
## **AMS Healthcare**

# (Erstwhile: Ace Medical Solutions (India) Pvt. Ltd.)

An ISO 9001-2015 Certified Company





GSTIN. 33AOSPA9153Q1ZA • DL No. TN-12-20B -00398, TN-12-21B-00398

## **CALIBRATION CERTIFICATE**

Customer	M/S. Bharathi Laboratory,	Certificate No:	AMS/HSP/289/5192
	#69/A, Swami Sannathi Street,	Issuing Date:	30/10/2023
Address:	Tirunelveli Town – 6.	Page No:	1 of 1

**Details of Device under Calibration (DUC):** 

DUC:	Micro Pipette	Id. No:	BH/LAB/FMP/002	
Make:		Department:	Laboratory	
Model:		Calibrated On:	28/10/2023	
Sl. No:	641400891	Calibration Due		
Type:	Fixed	Suggested:	27/10/2024	
Range &	100µl	DUC Condition		
Least Count:	•	on Receipt:	OK	

**Traceability Details:** 

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Electronic Weighing Scale	EWC/SF400C	PC/01/23/ATL/0512-02	31/01/2024
2	Thermo Hygrometer	Beetech/088	PC/01/23/ATL/0210-02	31/01/2024

Room Temperature: 25°C±2

Room Humidity: 30-60%

(At the time of Calibration)

### **Test Results:**

SL.	DUC Reading	Standard Reading (µl)		Correction
No.	(µl)	Trial 1	Trial 2	(µl)
1	100	100.00	100.00	0

Allowable Tolerance: User Defined

#### Note:

- 1. Calibration Points were selected as per customer request.
- 2. Each reading were taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings
- 3. Results Reported are valid at the time of and under the stated conditions of measurement

Calibrated By:

(Calibration Engineer)

ammilealthcare induse gmail.com

Approved by: (Technical Manager)

₫ 94434 74610, 93455 56148

aceagm@gmail.com