

CALIBRATION CERTIFICATE

| | | | |
|--------------------------|---|-----------------|------------------|
| Customer Name & Address: | M/S. Bharathi Laboratory, #69/A, Swami Sannathi Street, Tirunelveli Town – 6. | Certificate No: | AMS/HSP/289/5193 |
| | | Issuing Date: | 30/10/2023 |
| | | Page No: | 1 of 1 |

Details of Device under Calibration (DUC):

| | | | |
|----------------------|---------------|----------------------------|----------------|
| DUC: | Micro Pipette | Id. No: | BH/LAB/FMP/003 |
| Make: | --- | Department: | Laboratory |
| Model: | --- | Calibrated On: | 28/10/2023 |
| Sl. No: | 012217763 | Calibration Due Suggested: | 27/10/2024 |
| Type: | Fixed | DUC Condition on Receipt: | OK |
| Range & Least Count: | 500µl | | |

Traceability Details:

| Sl.No. | Test Equipment | Make | Certificate Number | Valid Up to |
|--------|---------------------------|-------------|----------------------|-------------|
| 1 | Electronic Weighing Scale | EWC/SF400C | PC/01/23/ATL/0512-02 | 31/01/2024 |
| 2 | Thermo Hygrometer | Beetech/088 | PC/01/23/ATL/0210-02 | 31/01/2024 |

Room Temperature: 25°C±2

Room Humidity: 30-60%

(At the time of Calibration)

Test Results:

| SL. No. | DUC Reading (µl) | Standard Reading (µl) | | Correction (µl) |
|---------|------------------|-----------------------|---------|-----------------|
| | | Trial 1 | Trial 2 | |
| 1 | 500 | 500.00 | 500.00 | 0 |

Allowable Tolerance: User Defined

Note:

1. Calibration Points were selected as per customer request.
2. Each reading were taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:
(Calibration Engineer)



aceagm@gmail.com



amshealthcaremdu@gmail.com

Approved by:
(Technical Manager)



94434 74610, 93455 56148