

AMS Healthcare

(Erstwhile : Ace Medical Solutions (India) Pvt. Ltd.)

An ISO 9001-2015 Certified Company

No. 770/1, Jothi Complex 2nd Floor, Alagar Kovil Main Road, Surveyor Colony,
Madurai - 625 007, Tamilnadu, India Mobile : +91 94434 74610



GSTIN. 33AOSPA9153Q1ZA • DL No. TN-12-20B -00398, TN-12-21B-00398

CALIBRATION CERTIFICATE

Customer Name & Address:	M/S. Bharathi Laboratory, #69/A, Swami Sannathi Street, Tirunelveli Town – 6.	Certificate No:	AMS/LAB/289/5189
		Issuing Date:	30/10/2023
		Page No:	1 of 1

Details of Device under Calibration (DUC):

DUC:	Centrifuge	Id. No:	BH/LAB/CEN/001
Make:	Remi	Department:	Laboratory
Model:	C-854/4	Calibrated On:	28/10/2023
Sl. No:	ZECN - 09466	Calibration Due Suggested:	27/10/2024
Type:	Manual	DUC Condition on Receipt:	OK
Range & Least Count:	0 – 3500 / LC - 1		

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Digital Tachometer	Mextech / DT-2234C	PC/01/23/ATL/0512-01	31/01/2024
2	Thermo Hygrometer	Beetech/088	PC/01/23/ATL/0210-02	31/01/2024

Room Temperature: 25°C±2 Room Humidity: 30-60%

(At the time of Calibration)

Test Results: Rotations Per Minute

SL.No.	DUC Reading (RPM)	Standard Reading (RPM)	Correction (RPM)
1	Position 1	1140	+0
2	Position 2	1680	+0
3	Position 3	2140	+0
4	Position 4	2640	+0
5	Position 5	3240	+0

Allowable Tolerance: Set by Customer

Note:

1. Calibration Points were selected as per customer request.
2. Each reading was taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:

(Calibration Engineer)



Approved by:

(Technical Manager)