

AMS Healthcare

(Erstwhile : Ace Medical Solutions (India) Pvt. Ltd.)

An ISO 9001-2015 Certified Company

No. 770/1, Jothi Complex 2nd Floor, Alagar Kovil Main Road, Surveyor Colony,
Madurai - 625 007, Tamilnadu, India Mobile : +91 94434 74610



GSTIN. 33AOSPA9153Q1ZA • DL No. TN-12-20B -00398, TN-12-21B-00398

CALIBRATION CERTIFICATE

Customer Name & Address:	M/S. Bharathi Laboratory, #69/A, Swami Sannathi Street, Tirunelveli Town – 6.	Certificate No:	AMS/LAB/289/5188
		Issuing Date:	30/10/2023
		Page No:	1 of 1

Details of Device under Calibration (DUC):

DUC:	Refrigerator – Reagent	Id. No:	BH/LAB/REFR/001
Make:	Whirlpool	Department:	Laboratory
Model:	FC 24 5S / 2010	Calibrated On:	28/10/2023
Sl. No:	---	Calibration Due Suggested:	27/10/2024
Type:	Reagent	DUC Condition on Receipt:	OK
Range & Least Count:	-5 to 8°C/ LC 1°C		

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	RTD Thermometer	Center / RTD 375	PC/01/23/ATL/0210-01	31/01/2024
2	Thermo Hygrometer	Beetech/088	PC/01/23/ATL/0210-01	31/01/2024

Room Temperature: 25°C±2

Room Humidity: 30-60%

(At the time of Calibration)

Test Results:

SL.No.	Place	DUC Reading (°C)	Standard Reading (°C)	Correction (°C)
1	Freezer	FRZ 1		
2		FRZ 2	-3.18	0
3			-2.96	0
4	Chiller	CHI 1	1.13	0
5		CHI 2	1.08	0
6	Chamber I	CHA 1	3.47	0
7		CHA 2	3.51	0
8	Chamber II	CHA 1	4.33	0
9		CHA 2	4.28	0
10	Chamber III	CHA 1	5.13	0
		CHA 2	5.24	0

Allowable Tolerance: User Defined

Note:

1. Calibration Points were selected as per customer request.
2. Each reading was taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:

(Calibration Engineer)

