

Lab Name Govt PHC, Manampatty

[Details About Robust Analysis](#)

Lab No 10534

[Detail About Monthly Summary](#)

Month May

[Detail about SDI](#)

Year 2023

Constituent Group Chemistry III

[Click on the analyte to view Graphical Data](#)

Date of Result Entered : 18/05/2023

Date of Report Published : 05/06/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1834	180.88	5.19	9.39	186 mg/dL	0.55	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1854	1.76	7.48	0.13	1.7 mg/dL	-0.45	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1798	114.60	5.08	5.83	111 mg/dL	-0.62	0.27

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

*S. N. J. Pr.*  
 ASSISTANT SURGEON  
 GOVT PRIMARY HEALTH CENTRE  
 MANAMPATTY - 693 405

Links

[Modify Address](#)

[Modify Methods](#)

[Submit Lab Results](#)

[View Lab Monthly Summary](#)