

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Manampathy

[Details About Robotal Analyze](#)

Lab No 10534

[Detail About Monthly Summary](#)

Month August

[Detail about BDI](#)

Year 2023

Constituent Group Chemistry III

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 19/08/2023


Date of Report Published : 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	BDI
						CV	SD		
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.85	0.13	0.00	189 mg/dL	1.49
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	0.77	0.12	1.6 mg/dL	-1.34
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.50	4.07	4.64	111 mg/dL	-0.60

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out


 ASSISTANT SURGEON
 GOVT. PRIMARY HEALTH CENTRE
 MANAMPATHY - 603 105.