

## CALIBRATION CERTIFICATE

Page no: 1 of 1

<b>Customer Name &amp; Address:</b> M/S GOVERNMENT PRIMARY HEALTH CENTRE, PEERKALAIKADU <sup>01</sup> , SIVAGANGAI - 630 303.		Certificate No:	VCPL/CC/23-24/32-002
		Date of Calibration:	03-10-2023
		Date of issue:	04-10-2023
		Date of Receipt:	03-10-2023
		Next cal Date:	02-10-2024
<b>Details of Device Under Calibration</b>			
Description:	SEMI AUTO ANALYZER	Make:	ROBONIK
Range:	-	Model/Type:	PRIETEST TOUCH
Resolution:	-	Condition of DUC	GOOD
Serial Number	ACTD0610321 RBK	Location:	LABORATORY
ID number	-	Cal. At:	ONSITE
<b>Environmental Conditions</b>		<b>Calibration Method Used</b>	
Temperature	25.2 °C	National/International standard:	IEC60601-1
Humidity	49 % Rh	Cal procedure no:	-

### Master Details

SL no:	Description	ID No / SL No	Certificate No	Make/Model	Valid Till
1	Electrical Safety Analyzer	6014025	ACCS230581	FLUKE/ESA-612	9-Mar-24


### Electrical Safety Report

S.NO	PARAMETERS	MEASURED VALUES $\mu$ A	ALLOWABLE LIMITS $\mu$ A
1	Earth Leakage	122.3	<5000 $\mu$ A For B,BF,CF
		<b>MEASURED VALUES <math>\mu</math>A</b>	<b>ALLOWABLE LIMITS <math>\mu</math>A</b>
2	Enclosure Leakage	187.6	<5000 $\mu$ A For B,BF,CF

**Remarks:**

- 1.DUC is Defined as Device Under Calibration.
- 2.Calibration Frequency decided by the customer.
- 3.The above instrument is working in good condition.

Calibrated By:

  
C.AKASH

(CALIBRATION ENGINEER)

Authorised by:

  
S. LAKSHMANAN MURUGAN,  
Technical Manager  
Vanavil Calibrations Pvt Ltd.

\*\*\* End of the Report \*\*\*

