

				CAL	IBRAT	ION	I CER	TIFIC	A7	Γ <u>E</u>		
CERTIFICATE NO: SBS/CL/23/17206						MEDICAL DEVICES					Page No:1 of 1	
Issue Date						04-11-2023						
SRF No & Date						SRF/23/01027-0001 & 03-11-2023						
Receipt Date						03-11-2023						
Calibration Date						03-11-2023						
Calibration Due						02-11-2024						
	omer Name	_										
GOV	ERNMENT	PRIMA	RY HEA	LTH CENTRE	Ξ,							
VALL	ANADU-62	1308,T	RICHY D	ISTRICT.								
•					Details of De	vice U	nder Calibra	tion (DUC	;}			
Descr	Description : SEMI AUTO ANALYZER				ER	Make & Model				ROBONIK & PRIETEST TOUCH		
Range : MULTI				Sr. No		:		ATCD1191220RBK				
Resolution : MULT		MULTI			Identification No		:		NA			
DUC Condition : SATISFACTORY			Location			:		LABORATORY				
				Environ	nental Condi	tions	& Calibration	Procedu	re D	etails		
Environmental Details Temperature:25.6 ° C					25.6 ° C	Relative Humidity				52% RH		
Calibration Procedure No SBS/CP/MD/2					Calibration		Calibration d	done at ONSITE		ONSITE		
					Refere	nce S	tandards De	tails				
S.No	Description Make/ SI N				Make/ SI No	o:		Certificate No				Validity
1	Electrical Safety Analyser F				Rigel Medical & 44L-1059			M-230809-16-4				10-08-2024

ELECTRICAL SAFETY RESULTS S.no **SPECIFICATION MEASURED VALUES EXPANDED UNCERTAINTY (±)** 1 Insulation Resisitance Measured values in MΩ Uncertainty in % (±) >20MΩ 13.92 2 Earth Leakage Measured values in µA Uncertainty in % (±) <5000µAfor B,BF,CF 3 **Enclosure Leakage** Measured values in µA Uncertainty in % (±) <500µAfor B,BF,CF 216 7.3

REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.

Equipment used for Calibration were calibrated & traceable to National & International Standards.
 Calibrated By

(Calibration Engineer)
M.DINESH





Chief Executive

Authorised Signatory