



Certificate of Calibration

Customer Name: Central Research Institute of Unani Medicine

Model: Fully Automated Bio Chemistry Analyzer EM-200

Serial No.: V200880

Calibration done date: 20/04/2023

Next Calibration due Date: On or Before 20/04/2024


Lab In charge: Dr. Mohd Naime

*This is to certify that above mentioned product has been verified of calibration
By*

*Checking Lamp Voltage
Checking gain of Photo meter for all Filters
Checking Cell blanks value
Checking laundry section
Checking movement of probe*

*Followed by calibration of all the three types of chemistries (picked up
CHOL, CREAT and SGPT) and QC run with satisfactory values of
QC*

Calibration at site performed by Mr. Sachin Singh


Authorized Signatory

Mr. Sachin Singh
Area Service Manager
Lucknow
For Transasia Bio-Medicals Ltd.

Result Reprint

Report Type : Calibration 12-Apr-2023 To 20-Apr-2023

Sr #	Lot #	Consumable	Test	Result Unit	Flag	Curve #	Result Date
1	11111	BLANK	SGPTD	0.0013 -		13	20-Apr-2023 15:28:12
2	2206133B	XL MULTICAL	SGPTD	-0.0334 -		15	20-Apr-2023 15:28:48
3	11111	BLANK	SGOTD	-0.0006 -		18	20-Apr-2023 15:29:42
4	11111	BLANK	SGOTD	-0.0007 -		19	20-Apr-2023 15:30:00
5	11111	BLANK	SGOTD	-0.0002 -		20	20-Apr-2023 15:30:18
6	2206133B	XL MULTICAL	SGOTD	-0.0363 -		21	20-Apr-2023 15:30:36
7	2206133B	XL MULTICAL	SGOTD	-0.0360 -		22	20-Apr-2023 15:30:54
8	2206133B	XL MULTICAL	SGOTD	-0.0381 -		23	20-Apr-2023 15:31:12
9	11111	BLANK	ALPU	0.0005 -		26	20-Apr-2023 15:32:06
10	11111	BLANK	ALPU	0.0006 -		27	20-Apr-2023 15:32:24
11	11111	BLANK	ALPU	0.0006 -		28	20-Apr-2023 15:32:42
12	2206133B	XL MULTICAL	ALPU	0.0690 -		29	20-Apr-2023 15:33:01
13	11111	BLANK	BIDD	-0.0008 -		34	20-Apr-2023 15:34:31
14	11111	BLANK	BIDD	0.0000 -		35	20-Apr-2023 15:34:49
15	11111	BLANK	BIDD	-0.0008 -		36	20-Apr-2023 15:35:07
16	2206133B	XL MULTICAL	BIDD	0.1035 -		37	20-Apr-2023 15:35:25
17	2206133B	XL MULTICAL	BIDD	0.1043 -		38	20-Apr-2023 15:35:43

Print Date

12-Sep-2023

Result Reprint

Report Type : Calibration 12-Apr-2023 To 20-Apr-2023

Sr #	Lot #	Consumable	Test	Result Unit	Flag	Curve #	Result Date
18	2206133B	XL MULTICAL	BIDD	0.1037 -		39	20-Apr-2023 15:36:01
19	111111	BLANK	BIT	0.0006 -		42	20-Apr-2023 15:36:55
20	111111	BLANK	BIT	-0.0009 -		43	20-Apr-2023 15:37:13
21	111111	BLANK	BIT	-0.0001 -		44	20-Apr-2023 15:37:31
22	2206133B	XL MULTICAL	BIT	0.1840 -		45	20-Apr-2023 15:37:49
23	2206133B	XL MULTICAL	BIT	0.1835 -		46	20-Apr-2023 15:38:08
24	2206133B	XL MULTICAL	BIT	0.1842 -		47	20-Apr-2023 15:38:26
25	111111	BLANK	UREA	-0.0019 -		52	20-Apr-2023 15:39:56
26	111111	BLANK	UREA	-0.0030 -		53	20-Apr-2023 15:40:14
27	111111	BLANK	UREA	-0.0073 -		54	20-Apr-2023 15:40:32
28	2206133B	XL MULTICAL	UREA	-0.1324 -		55	20-Apr-2023 15:40:50
29	2206133B	XL MULTICAL	UREA	-0.1300 -		56	20-Apr-2023 15:41:08
30	2206133B	XL MULTICAL	UREA	-0.1325 -		57	20-Apr-2023 15:41:26
31	111111	BLANK	GLU	0.0102 -		60	20-Apr-2023 15:44:45
32	111111	BLANK	GLU	0.0114 -		61	20-Apr-2023 15:45:03
33	111111	BLANK	GLU	0.0097 -		62	20-Apr-2023 15:45:21
34	2206133B	XL MULTICAL	GLU	0.7663 -		63	20-Apr-2023 15:45:39

Result Reprint

Report Type : Calibration 12-Apr-2023 To 20-Apr-2023

Sr #	Lot #	Consumable	Test	Result Unit	Flag	Curve #	Result Date
35	2206133B	XL MULTICAL	GLU	0.6793 -		64	20-Apr-2023 15:45:57
36	2206133B	XL MULTICAL	GLU	0.7697 -		65	20-Apr-2023 15:46:15
37	11111	BLANK	CHOL	0.0294 -		68	20-Apr-2023 15:47:09
38	11111	BLANK	CHOL	0.0273 -		69	20-Apr-2023 15:47:27
39	11111	BLANK	CHOL	0.0297 -		70	20-Apr-2023 15:47:45
40	2206133B	XL MULTICAL	CHOL	0.2831 -		71	20-Apr-2023 15:48:03
41	2206133B	XL MULTICAL	CHOL	0.2827 -		72	20-Apr-2023 15:48:21
42	2206133B	XL MULTICAL	CHOL	0.2840 -		73	20-Apr-2023 15:48:39
43	11111	BLANK	UA	0.0055 -		76	20-Apr-2023 15:49:33
44	11111	BLANK	UA	0.0056 -		77	20-Apr-2023 15:49:52
45	11111	BLANK	UA	0.0049 -		78	20-Apr-2023 15:50:10
46	2206133B	XL MULTICAL	UA	0.0816 -		79	20-Apr-2023 15:50:27
47	2206133B	XL MULTICAL	UA	0.0819 -		80	20-Apr-2023 15:50:46
48	2206133B	XL MULTICAL	UA	0.0809 -		81	20-Apr-2023 15:51:04
49	11111	BLANK	CA	0.2829 -		84	20-Apr-2023 15:51:58
50	11111	BLANK	CA	0.2891 -		85	20-Apr-2023 15:52:16
51	11111	BLANK	CA	0.2923 -		86	20-Apr-2023 15:52:34

Result Reprint

Report Type : Calibration 12-Apr-2023 To 20-Apr-2023

Sr #	Lot #	Consumable	Test	Result Unit	Flag	Curve #	Result Date
52	2206133B	XL MULTICAL	CA	0.6351 -		87	20-Apr-2023 15:52:52
53	2206133B	XL MULTICAL	CA	0.6441 -		88	20-Apr-2023 15:53:10
54	2206133B	XL MULTICAL	CA	0.6461 -		89	20-Apr-2023 15:53:28
55	11111	BLANK	ALBD	0.0755 -		92	20-Apr-2023 15:54:22
56	11111	BLANK	ALBD	0.0775 -		93	20-Apr-2023 15:54:40
57	11111	BLANK	ALBD	0.0776 -		94	20-Apr-2023 15:54:58
58	2206133B	XL MULTICAL	ALBD	0.7109 -		95	20-Apr-2023 15:55:16
59	2206133B	XL MULTICAL	ALBD	0.7120 -		96	20-Apr-2023 15:55:35
60	2206133B	XL MULTICAL	ALBD	0.7032 -		97	20-Apr-2023 15:55:52
61	11111	BLANK	PRO	0.0955 -		100	20-Apr-2023 15:56:46
62	11111	BLANK	PRO	0.0969 -		101	20-Apr-2023 15:57:04
63	11111	BLANK	PRO	0.0967 -		102	20-Apr-2023 15:57:23
64	2206133B	XL MULTICAL	PRO	0.4313 -		103	20-Apr-2023 15:57:41
65	2206133B	XL MULTICAL	PRO	0.4132 -		104	20-Apr-2023 15:57:59
66	2206133B	XL MULTICAL	PRO	0.4138 -		105	20-Apr-2023 15:58:17
67	11111	BLANK	CRE1	0.0069 -		108	20-Apr-2023 15:59:11
68	11111	BLANK	CRE1	0.0069 -		109	20-Apr-2023 15:59:29

Result Reprint

Report Type : Calibration 12-Apr-2023 To 20-Apr-2023

Sr #	Lot #	Consumable	Test	Result Unit	Flag	Curve #	Result Date
69	11111	BLANK	CRE1	0.0078	-	110	20-Apr-2023 15:59:47
70	2206133B	XL MULTICAL	CRE1	0.0880	-	111	20-Apr-2023 16:00:05
71	2206133B	XL MULTICAL	CRE1	0.0828	-	112	20-Apr-2023 16:00:23
72	2206133B	XL MULTICAL	CRE1	0.0827	-	113	20-Apr-2023 16:00:41