REMI GROUP

he Product

09.8

WARRANTY CARD

Product Name: Remi Entrifique
Model No.:
Sr. No.: 270N-16370
Date of Purchase:
Supplier's Name: Gryan Enterprises, Lal Rogh, Luckhow
Address:
Bill / Cash Memo No.: GST 1999 123-24 06-06-202

Packed By

REMI ELEKTROTECHNIK LTD.

eeked By

(INSTRUMENTS DIVISION).

Building: "B" Survey No. 65/1, Village, Valiv Vasai (E),

Dist. Palghar - 401 208, INDIA

Tel.: +91 844 601 31 82 to 87

Approved By

Email: sales@remilabworld.com

Web: www.remilabworld.com



Calibration Certificate

				$\overline{}$
UNIT DESCRIPTION:	CLINICAL CENTRI	FUGE MODEL N	O.: R - 303	
SR. NO.:	ZIDN-16370	CERTIFICATE N	IO: RMTC - 0959/22	
NAME OF CUSTOMER:				
ADDRESS OF CUSTOMER:				
CALIBRATION DATE:	12.04.2023	CAL. VALIDI	ΓY: 11.04.2024	
DETAILS OF EQUIPMENT UNOTE: Tick I mark where		TION		
Digital Lab Thermometer	Di	gital Tachometer	Digital Stop Watch	
Auto Ranging R.P.M. Meter	Di	gital Strobe	Weight Set	
Digital Temperature Indicato	or		_	

Sr. No.	Description	Certified By	Instrument Id. No.	Certificate No.	Calibration Date	Validity
01	Digi. Strobe	Autocal	ISR 01	C21101836/P/01	27.10.2022	26.10.2023

OBSERVATIONS:

Sr. No.	Speed Reading on Digital Strobe	Remarks
01	1004	OK
03	2514	OK
05	3818	OK

NOTE: 1. Calibration Traceable to national standards.

- 2. Overall deviation is within specified accuracy.
- 3. Result is related only to the item calibrated.
- 4. Maximum tolerance permissible is ± 1°C & ± 5°C whichever is applicable and RPM ±2% FSD and ±5gm to set weight.

REMARKS: ACCEPTED

CALIBRATED BY: VIJAY KADAM

APPROVED BY: MANAGER Q. A.

Doc. No.: QA/CC/F, Rev:00, Date:02.01.2010