AMS Healthcare

(Erstwhile: Ace Medical Solutions (India) Pvt. Ltd.)

An ISO 9001-2015 Certified Company

No. 770/1, Jothi Complex 2nd Floor, Alagar Kovil Main Road, Surveyor Colony, Madurai - 625 007, Tamilnadu, India Mobile : +91 94434 74610



GSTIN. 33AOSPA9153Q1ZA • DL No. TN-12-20B -00398, TN-12-21B-00398

CALIBRATION CERTIFICATE

				1
ſ	Customer	M/S. SRI Computerised Laboratory,	Certificate No:	AMS/LAB/289/5199
1		#36B, South Car Street, Vagaiyadi Corner,	Issuing Date:	30/10/2023
1	Address:	Tirunelyeli Town -6.	Page No:	1 of 1
- 1	Audicss.	Thunciven Town 6.		

Details of Device under Calibration (DUC):

Details of Device under Calibration (DUC):				
DUC:	Micropipette	Id. No:	SRI/LAB/VMP/001	
Make:	Dragon Lab	Department:	Laboratory	
Model:		Calibrated On:	28/10/2023	
Sl. No:	YE5A426047	Calibration		
Type:	Variable	Due Suggested:	27/10/2024	
Range &	5 to 50μl	DUC Condition		
Least Count:		on Receipt:	OK	

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Electronic Weighing Scale	EWC/SF400C	PC/01/23/ATL/0512-02	31/01/2024
2	Thermo Hygrometer	Beetech/088	PC/01/23/ATL/0210-01	31/01/2024

Room Temperature: 25°C±2

Room Humidity: 30-60%

(At the time of Calibration)

Test ResultsVolume µl:

SL.	DUC Reading	Reading Standard Reading (µl)		Correction	
No.	(μl)	Trial 1	Trial 2	(µl)	
1	05	5.0	5.0	+0.0	
2	10	10.0	10.0	+0.0	
3	20	20.0	20.0	+0.0	
1	30	30.0	30.0	+0.0	
5	50	50.0	50.0	+0.0	

Allowable Tolerance: User Defined

Note:

1. Calibration Points were selected as per customer request.

2. Each reading was taken by Repeatability Test(i.e.) each reading in a row is a mean of 101

3. Results Reported are valid at the time of and underthe stated conditions of measurements.

Calibrated By:

(Calibration Engineer)

amske Missakehodu@gmail.com

Approved by: (Technical Manager)

A

94434 74610, 93455 56148

aceagm@gmail.com