



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - SEPTEMBER 2023



Lab Name: Govt. Block PHC, Veeramachanpatti

Constituent Group: Chemistry III

PT stem: Lyophilized human serum based

Lab No: 7245

Date of Result Entered: 20/09/2023

Date of Report Published: 05/10/2023

Sl. No.	Test Name	Method	Result	Reference Range	SDI	U
1	GLUCOSE III	GOD-POD III	1870	310.86 - 482	14.98	332 mg/dL 1.41
2	UREA III	Urease UV / GLDH III	187	35.19 - 12.73	4.88	30 mg/dL -1.16
3	CREATININE III	Jaffe's Kinetic - Alkaline Picrate III	1852	1.29 - 0.14	0.10	1.4 mg/dL 1.05
4	CHOLESTEROL III	CHOD-PAP III	1783	113.62 - 3.80	4.32	118 mg/dL 1.93
SDI Range			Interpretation			
Within -1.00 to +1.00			Excellent.			
Within ±1.01 to ±2.00			Good.			
Within ±2.01 to ±2.99			Accept with caution. Warning Signal.			
Beyond ±3.0			Unacceptable performance. Action Signal.			

LAB ADDRESS
Govt. Block PHC, Veeramachanpatti,
THE BLOCK MEDICAL OFFICER, TRICHY DISTRICT
Veeramachanpatti,
TAMIL NADU 621206

Coordinator Contact Details
E-mail: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Homogeneity and Stability of the sample is passed
Data in CMC EQAS reports is confidential
CMC EQAS does not sub-contract any components
..... End of Report

Dr. Pamela Christodoss
CMC EQAS Coordinator
Christian Medical College, Vellore

MEDICAL OFFICER
GOVT. PRIMARY HEALTH CENTRE
VEERAMACHANPATTI,
THURAIYUR.