



CENTRE for HIV TESTING (CHT)  
State Reference Laboratory for HIV  
Department of Microbiology  
Lady Hardinge Medical College & Associated Hospitals  
New Delhi-110001

**THERMOMETER CALIBRATION CERTIFICATE OF GLM, ICTC & PPTCT**

Name of the thermometer to be calibrated: glass thermometers 01

ID of the thermometer to be calibrated: GLM/Thermometers/2023/01

Temperature range of the thermometer: -10°C to +110°C

Name & ID of reference thermometer: <sup>LHMC</sup> SRL/G. Thermometers/2023/35

Temperature range of reference thermometer: -10°C to +110°C

Date of calibration: 07/11/23

Time of reading	Temperature of working thermometer	Temperature of reference thermometer	Difference in readings
After 1 hour	4°C	4°C	0
After 2 hours	4°C	4°C	0
After 3 hours	4°C	4°C	0

Correction factor: 0 (zero)

Remarks: NA

Calibration done by: Sandhya mehta, LT, SRL, LHMC, SRL

Verified by: Rajalini Nayak *Rajalini* <sup>Test. officer, SRL LHMC</sup>

Approved by: Dr. Sanjib Gogoi *Dr. Sanjib Gogoi* <sup>Nodal officer, SRL CHT, LHMC</sup>