



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMCEXTERNALQUALITYASSURANCESHEME

MONTHLY SUMMARY REPORT - OCTOBER 2023



PC-1024

Lab Name Govt. UPHC, Subramaniyar Street  
Saidapet

Lab No

9295

Constituent Group Chemistry III

Date of Result Entered :

19/10/2023

PT item Lyophilized human serum based

Date of Report Published :

04/11/2023

Sl.No.	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD/POD III	Robonick ( Semi automatic / Fully Automatic)	1905	118.63	6.44	7.63	120.7 mg/dL	0.27	0.35
2	CREATININE III	GOD/POD III	Robonick ( Semi automatic / Fully Automatic)	1912	7.12	5.02	0.36	7.2 mg/dl	-0.19	0.02
3	CHOLETEROL III	GOD/POD III	Robonick ( Semi automatic / Fully Automatic)	1786	111.59	4.28	4.78	104.3 mg/dL	-1.53	0.23

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within +1.01 to + 2.00	Good.
Within +2.01 to +2.99	Accept with caution. Warning Signal.
Beyond +: 3.0	Unacceptable performan"ce. Action Signal.

**LAB ADDRESS :**

Govt. UPHC, Suburamaniyar Street, Saidapet,  
SAIDAPET, SUBURAMAIYAR STREET  
Subramaniyar Street , Saidapet  
TAMILNADU632012

*Pamela Christudoss*

Coordinator Contact Details:  
Email: clinqc@cmcvellore.ac.in  
contactNumber: 0416-2283102

Dr. Pamelachristudoss  
CMCEQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMCEQAS does not sub contract any components  
\*\*\*\*\*End of Report\*\*\*\*\*