



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMCEXTERNALQUALITYASSURANCESHEME

MONTHLY SUMMARY REPORT - SEPTEMBER 2023



PC-1024

Lab Name Govt. UPHC, Subramaniyar Street  
Saidapet

Lab No

9295

Constituent Group Chemistry III

Date of Result Entered :

19/09/2023

PT item Lyophilized human serum based

Date of Report Published :

05/10/2023

Sl. No.	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD/POD III	Robonick ( Semi automatic / Fully Automatic)	1870	310.86	4.82	14.98	309.6 mg/dl	-0.89	0.42
2	CREATININE III	GOD/POD III	Robonick ( Semi automatic / Fully Automatic)	1852	1.29	8.14	0.10	1.2 mg/dl	-0.17	0.01
3	CHOLESTEROL III	GOD/POD III	Robonick ( Semi automatic / Fully Automatic)	1783	113.62	3.80	4.32	112 mg/dl	-0.39	0.22

SDI Range	Interpretation
Within -1.00 to $\pm$ 1.00	Excellent.
Within $\pm$ 1.01 to $\pm$ 2.00	Good.
Within $\pm$ 2.01 to $\pm$ 2.99	Accept with caution. Warning Signal.
Beyond $\pm$ 3.0	Unacceptable performance. Action Signal.

**LAB ADDRESS :**

Govt. UPHC, Suburamaniyar Street, Saidapet,  
SAIDAPET, SUBURAMAIYAR STREET  
Subramaniyar Street , Saidapet  
TAMILNADU 632012

*Pamela Christudoss*

Coordinator Contact Details:  
Email: clinqc@cmcvellore.ac.in  
contactNumber: 0416-2283102

Dr. Pamela Christudoss  
CMCEQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not subcontract any components  
\*\*\*\*\*End of Report\*\*\*\*\*