



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMCEXTERNALQUALITYASSURANCEScheme



PC-1024

MONTHLY SUMMARY REPORT - AUGUST 2023

Lab Name Govt. UPHC, Subramaniyar Street
Saidapet

Lab No

9295

Constituent Group Chemistry III

Date of Result Entered :

19/08/2023

PT item Lyophilized human serum based

Date of Report Published :

07/09/2023

Sl.No.	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD/POD III	Robonick (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	167.6 mg/dl	-0.89	0.42
2	CREATININE III	GOD/POD III	Robonick (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.74 mg/dl	-0.17	0.01
3	CHOLETEROL III	GOD/POD III	Robonick (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	112 mg/dl	-0.39	0.22

SDI Range	Interpretation
Within -1.00 to \pm 1.00	Excellent.
Within \pm 1.01 to \pm 2.00	Good.
Within \pm 2.01 to \pm 2.99	Accept with caution. Warning Signal.
Beyond \pm 3.0	Unacceptable performan"ce. Action Signal.

LAB ADDRESS :

Govt. UPHC, Suburamaniyar Street, Saidapet,
SAIDAPET, SUBURAMAIYAR STREET
Subramaniyar Street , Saidapet
TAMILNADU632012

Pamela Christudoss

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
contactNumber: 0416-2283102

Dr.Pamelachristudoss
CMCEQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not subcontract any components

*****End of Report*****