



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMCEXTERNALQUALITYASSURANCESHEME

MONTHLY SUMMARY REPORT -JULY 2023



PC-1024

9295

Lab Name Govt. UPHC, Subramaniyar Street
Saidapet

Lab No

Constituent Group Chemistry III

Date of Result Entered :

19/07/2023

PT item Lyophilized human serum based

Date of Report Published :

07/08/2023

Sl. No.	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U.
						CV	SD			
1	GLUCOSE III	GOD/POD III	Robonick (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	110 mg/dl	0.11	0.28
2	CREATININE III	GOD/POD III	Robonick (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	5.5mg	-0.94	0.02
3	CHOLESTEROL III	GOD/POD III	Robonick (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	112 mg/dl	-0.21	0.22

SDI Range	Interpretation
Within -1.00 to ± 1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. UPHC, Subramaniyar Street, Saidapet,
SAIDAPET, SUBURAMAIYAR STREET
Subramaniyar Street , Saidapet
TAMILNADU 632012

Pamela Christudoss

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
contactNumber: 0416-2283102

Dr. Pamela Christudoss
CMCEQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
*****End of Report*****