


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO.	: 230617B/N-112	Page	: 1 of 1	
CALIBRATED ON.	: 18 / 06 / 2023	Ref. Std.	: IS:12508:1988	
CALIBRATION DUE.	: 17 / 06 / 2024	C.P. No.	: ACS/CP/10	
SRF No. & DATE	: 230617B/N Dt. 17.06.2023	Phy. Inspection	: O.K.	
ULR No. & ISSUE DATE	: CC342722000000520F Dt. 21.06.2023	Cal. At	: Site / Lab	
CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt. Ltd. 126 Lajpat Nagar Block 10-A1, New Nakodhar Chowk, Jalandhar Punjab-144001			
ITEM CALIBRATED	: Centrifuge			
	Make	: Remi	Range	: (0 to 6000) RPM
	L. Count	:	ID No.	: RCL/CP/Cent-02
	Location	: Clinical Pathology	S. No.	: ZIFN-27241
ENVIRONMENT CONDITION:	Temp. { 25°C + 3°C} R.H. {40% to 60%}			
STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Tachometer	ACS/DTM/01/ 211-0155V	Godrej	C-220912-6-1	12/09/2023
RESULTS	: Mechanical Discipline			
	* U. C. INSTT. READS (RPM)	Standard Measured (RPM)	Error (RPM)	
	1000	1001.32	- 1.32	
	2000	2002.46	- 2.46	
	4000	4005.78	- 5.78	
	6000	6006.84	- 6.84	
	(average of readings)			

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 5.0 rpm of reading at coverage Factor $k=1.96$.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

