

ADVANCE CALIBRATION SERVICES

(ISO: 17025 NABL ACCREDITED LAB.)



'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob.: +91-9891121066, 9868309579 • E-mail: advance.cal@gmail.com

Calibration Certificate

CERTIFICATE NO.	: 230617B/N-118	Page	: 1 of 1
CALIBRATED ON.	: 18 / 06 / 2023	Ref. Std.	: CG-11
CALIBRATION DUE.	: 17 / 06 / 2024	C.P. No.	: ACS/CP/07
SRF No. & DATE	: 230617B/N Dt. 17.06.2023	Phy. Inspec	tion : O.K.
ULR No. & ISSUE DATE	: CC342722000000526F Dt. 21.06.2023	Cal. At	: Site / Lab
CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt. Ltd. 126 Lajpat Nagar Block 10-A1, New I Jalandhar Punjab-144001	Nakodhar Chow	/ k ,

ITEM CALIBRATED	: Digital Tem	perature Indicator		
	Make	: RoHS	Range	: (-50 to 110) °C
	L. Count	:	ID No.	: RCL/SR/D.Temp-06
	Location	: Sample Receiving Dept.	S. No.	: N.A

ENVIRONMENT CONDITION:	Temp.	{ 25°C <u>+</u> 3°C}	R.H.	{40% to 60%}
------------------------	-------	----------------------	------	--------------

STANDAKD INSTI	RUMENTS USED:			
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Digital Calibrator	ACS/UC/01 / 300561	Anshaankan	AIPL/2022/0846	10/11/2023

RESULTS: ET Discipline		6 1
* U. C. INSTT.	STANDARD	ERROR
READS (°C)	READING (°C)	(°C)
2.06	2	0.06
4.20	4	0.20
6.36	6	0.36
8.42	8	0.42
(3	average of readings)	

REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is ± 1.0°C at coverage Factor k=1.96.

Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.

 The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above. This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use
- Confirmity statement not provided since not required.



